

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
 Data File : VN079910.D
 Acq On : 04 Nov 2023 22:03
 Operator : JC\MD
 Sample : 05234-05
 Misc : 1.12g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ATH-OILY-DEBRIS

Quant Time: Nov 06 02:22:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	242893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	470442	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	500932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	224047	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	211496	57.152	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	114.300%	
35) Dibromofluoromethane	8.165	113	172074	51.680	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	103.360%	
50) Toluene-d8	10.571	98	627141	51.164	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.320%	
62) 4-Bromofluorobenzene	12.853	95	233104	57.342	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	114.680%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	498688	383.648	ug/l	94
20) Methylene Chloride	5.259	84	4350	1.326	ug/l #	91
52) Toluene	10.630	92	9850	1.167	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
Data File : VN079910.D
Acq On : 04 Nov 2023 22:03
Operator : JC\MD
Sample : 05234-05
Misc : 1.12g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ATH-OILY-DEBRIS

Quant Time: Nov 06 02:22:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

