

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
 Data File : VN079889.D
 Acq On : 04 Nov 2023 13:19
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 06 02:11:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.602	0.649	-7.8	91	0.00
3 P	Chloromethane	0.837	0.726	13.3	83	0.00
4 C	Vinyl Chloride	0.788	0.963	-22.2#	104	0.00
5 T	Bromomethane	0.479	0.569	-18.8	107	0.00
6 T	Chloroethane	0.623	0.587	5.8	85	0.00
7 T	Trichlorofluoromethane	0.981	1.016	-3.6	90	0.00
8 T	Diethyl Ether	0.372	0.402	-8.1	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.568	0.603	-6.2	94	0.00
10 T	Methyl Iodide	0.649	0.699	-7.7	87	0.00
11 T	Tert butyl alcohol	0.097	0.091	6.2	84	0.00
12 CM	1,1-Dichloroethene	0.538	0.556	-3.3#	88	0.00
13 T	Acrolein	0.080	0.095	-18.8	105	0.00
14 T	Allyl chloride	0.824	0.921	-11.8	97	0.00
15 T	Acrylonitrile	0.303	0.340	-12.2	98	0.00
16 T	Acetone	0.268	0.329	-22.8	113	0.00
17 T	Carbon Disulfide	1.552	1.532	1.3	87	0.00
18 T	Methyl Acetate	1.280	1.422	-11.1	101	0.00
19 T	Methyl tert-butyl Ether	1.853	2.087	-12.6	95	0.00
20 T	Methylene Chloride	0.675	0.706	-4.6	95	0.01
21 T	trans-1,2-Dichloroethene	0.600	0.639	-6.5	93	0.00
22 T	Diisopropyl ether	1.910	2.346	-22.8	101	0.00
23 T	Vinyl Acetate	1.054	1.225	-16.2	94	0.00
24 P	1,1-Dichloroethane	1.161	1.334	-14.9	100	0.00
25 T	2-Butanone	0.392	0.477	-21.7	103	0.00
26 T	2,2-Dichloropropane	0.960	1.040	-8.3	91	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.790	-10.0	96	0.00
28 T	Bromochloromethane	0.566	0.643	-13.6	97	0.00
29 T	Tetrahydrofuran	0.255	0.297	-16.5	99	0.00
30 C	Chloroform	1.261	1.359	-7.8#	99	0.00
31 T	Cyclohexane	1.007	1.064	-5.7	96	0.00
32 T	1,1,1-Trichloroethane	1.028	1.152	-12.1	98	0.00
33 S	1,2-Dichloroethane-d4	0.762	0.858	-12.6	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	86	0.00
35 S	Dibromofluoromethane	0.354	0.383	-8.2	92	0.00
36 T	1,1-Dichloropropene	0.505	0.572	-13.3	98	0.00
37 T	Ethyl Acetate	0.463	0.541	-16.8	98	0.00
38 T	Carbon Tetrachloride	0.509	0.575	-13.0	97	0.00
39 T	Methylcyclohexane	0.476	0.533	-12.0	91	0.00
40 TM	Benzene	1.513	1.693	-11.9	98	0.00
41 T	Methacrylonitrile	0.246	0.303	-23.2	104	0.00
42 TM	1,2-Dichloroethane	0.529	0.634	-19.8	105	0.00
43 T	Isopropyl Acetate	0.813	0.888	-9.2	95	0.00
44 TM	Trichloroethene	0.376	0.405	-7.7	95	0.00
45 C	1,2-Dichloropropane	0.387	0.466	-20.4#	103	0.00
46 T	Dibromomethane	0.264	0.306	-15.9	100	0.00
47 T	Bromodichloromethane	0.563	0.630	-11.9	100	0.00
48 T	Methyl methacrylate	0.356	0.440	-23.6	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
 Data File : VN079889.D
 Acq On : 04 Nov 2023 13:19
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 06 02:11:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	92	0.00
50 S	Toluene-d8	1.303	1.404	-7.8	89	0.00
51 T	4-Methyl-2-Pentanone	0.458	0.557	-21.6	100	0.00
52 CM	Toluene	0.897	1.045	-16.5#	96	0.00
53 T	t-1,3-Dichloropropene	0.552	0.655	-18.7	96	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.710	-17.9	95	0.00
55 T	1,1,2-Trichloroethane	0.361	0.411	-13.9	97	0.00
56 T	Ethyl methacrylate	0.515	0.629	-22.1	96	0.00
57 T	1,3-Dichloropropane	0.636	0.745	-17.1	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.261	0.291	-11.5	88	0.00
59 T	2-Hexanone	0.350	0.407	-16.3	96	0.00
60 T	Dibromochloromethane	0.386	0.451	-16.8	96	0.00
61 T	1,2-Dibromoethane	0.362	0.415	-14.6	96	0.00
62 S	4-Bromofluorobenzene	0.432	0.488	-13.0	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.324	0.399	-23.1	107	0.00
65 PM	Chlorobenzene	1.085	1.171	-7.9	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.442	-13.0	97	0.00
67 C	Ethyl Benzene	1.890	2.142	-13.3#	94	0.00
68 T	m/p-Xylenes	0.704	0.805	-14.3	95	0.00
69 T	o-Xylene	0.678	0.766	-13.0	93	0.00
70 T	Styrene	1.089	1.321	-21.3	94	0.00
71 P	Bromoform	0.277	0.326	-17.7	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.795	4.186	-10.3	94	0.00
74 T	N-amyl acetate	1.656	1.723	-4.0	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.345	1.364	-1.4	93	0.00
76 T	1,2,3-Trichloropropane	1.174	1.263	-7.6	99	0.00
77 T	Bromobenzene	0.929	0.999	-7.5	95	0.00
78 T	n-propylbenzene	4.346	5.022	-15.6	95	0.00
79 T	2-Chlorotoluene	2.776	3.090	-11.3	98	0.00
80 T	1,3,5-Trimethylbenzene	2.952	3.463	-17.3	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.431	0.481	-11.6	89	0.00
82 T	4-Chlorotoluene	2.759	3.074	-11.4	96	0.00
83 T	tert-Butylbenzene	2.417	2.778	-14.9	94	0.00
84 T	1,2,4-Trimethylbenzene	2.905	3.460	-19.1	95	0.00
85 T	sec-Butylbenzene	3.348	3.871	-15.6	94	0.00
86 T	p-Isopropyltoluene	2.730	3.234	-18.5	94	0.00
87 T	1,3-Dichlorobenzene	1.728	1.869	-8.2	95	0.00
88 T	1,4-Dichlorobenzene	1.754	1.850	-5.5	95	0.00
89 T	n-Butylbenzene	2.351	2.745	-16.8	93	0.00
90 T	Hexachloroethane	0.541	0.589	-8.9	100	0.00
91 T	1,2-Dichlorobenzene	1.632	1.776	-8.8	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.228	0.253	-11.0	93	0.00
93 T	1,2,4-Trichlorobenzene	0.776	0.851	-9.7	90	0.00
94 T	Hexachlorobutadiene	0.352	0.391	-11.1	100	0.00
95 T	Naphthalene	2.631	2.601	1.1	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110423\
Data File : VN079889.D
Acq On : 04 Nov 2023 13:19
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 06 02:11:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.760	0.820	-7.9	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6