

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110518\
 Data File : VN052152.D
 Acq On : 5 Nov 2018 17:40
 Operator : MD\SY
 Sample : VN1105WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1105WBL01

Quant Time: Nov 06 05:54:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 22:32:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	354510	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	617301	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	600227	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	251058	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	252693	60.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.22%	
35) Dibromofluoromethane	7.59	113	229388	59.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	119.70%	
50) Toluene-d8	10.09	98	914717	61.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	123.46%	
62) 4-Bromofluorobenzene	12.40	95	305720	62.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	125.70%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110518\
Data File : VN052152.D
Acq On : 5 Nov 2018 17:40
Operator : MD\SY
Sample : VN1105WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN1105WBL01

Quant Time: Nov 06 05:54:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 22:32:25 2018
Response via : Initial Calibration

