

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110518\
 Data File : VN052156.D
 Acq On : 5 Nov 2018 19:27
 Operator : MD\SY
 Sample : J5812-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP2-B

Manual Integrations
 APPROVED

MMDadoda
 11/6/2018 10:00:25 AM

Quant Time: Nov 06 06:04:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 22:32:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	347059	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	623990	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	598825	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	254143	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	264495	64.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.54%	
35) Dibromofluoromethane	7.59	113	229305	59.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.38%	
50) Toluene-d8	10.09	98	928849	62.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.02%	
62) 4-Bromofluorobenzene	12.40	95	312429	63.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.08%	
Target Compounds						
16) Acetone	3.81	43	8311m	8.914	ug/l	
18) Methyl Acetate	4.33	43	4108m	1.386	ug/l	
20) Methylene Chloride	4.56	84	154040	39.761	ug/l	94
43) Isopropyl Acetate	8.17	43	203942	32.380	ug/l #	89
95) Naphthalene	15.13	128	10728	1.378	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110518\
Data File : VN052156.D
Acq On : 5 Nov 2018 19:27
Operator : MD\SY
Sample : J5812-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP2-B

Manual Integrations
APPROVED

MMDadoda
11/6/2018 10:00:25 AM

Quant Time: Nov 06 06:04:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 22:32:25 2018
Response via : Initial Calibration

