

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110518\
 Data File : VN052159.D
 Acq On : 5 Nov 2018 20:47
 Operator : MD\SY
 Sample : J5812-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP2-A

Quant Time: Nov 06 06:11:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 22:32:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	343505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	614508	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	601334	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	253132	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	260066	63.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	127.70%	
35) Dibromofluoromethane	7.59	113	230163	60.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.66%	
50) Toluene-d8	10.09	98	913086	61.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.80%	
62) 4-Bromofluorobenzene	12.40	95	310789	64.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.36%	
Target Compounds						
16) Acetone	3.82	43	7114	7.709	ug/l	99
20) Methylene Chloride	4.55	84	23720	6.186	ug/l	94
43) Isopropyl Acetate	8.17	43	200812	32.375	ug/l #	88
95) Naphthalene	15.13	128	23331	3.010	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110518\
Data File : VN052159.D
Acq On : 5 Nov 2018 20:47
Operator : MD\SY
Sample : J5812-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP2-A

Quant Time: Nov 06 06:11:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 22:32:25 2018
Response via : Initial Calibration

