

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110518\
 Data File : VN052161.D
 Acq On : 5 Nov 2018 21:40
 Operator : MD\SY
 Sample : J5812-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP3-A

Manual Integrations
 APPROVED

MMDadoda
 11/6/2018 10:00:31 AM

Quant Time: Nov 06 06:15:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 22:32:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	332171	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	601015	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	573152	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	238861	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	256844	65.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	130.42%	
35) Dibromofluoromethane	7.59	113	226248	60.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.26%	
50) Toluene-d8	10.09	98	888030	61.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.10%	
62) 4-Bromofluorobenzene	12.40	95	299092	63.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	126.32%	
Target Compounds						
16) Acetone	3.82	43	10298	11.540	ug/l	99
20) Methylene Chloride	4.56	84	12037m	3.246	ug/l	
43) Isopropyl Acetate	8.17	43	197134	32.496	ug/l #	88
95) Naphthalene	15.13	128	314432	42.988	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110518\
Data File : VN052161.D
Acq On : 5 Nov 2018 21:40
Operator : MD\SY
Sample : J5812-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP3-A

Manual Integrations
APPROVED

MMDadoda
11/6/2018 10:00:31 AM

Quant Time: Nov 06 06:15:30 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 22:32:25 2018
Response via : Initial Calibration

