

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110518\
 Data File : VN052162.D
 Acq On : 5 Nov 2018 22:07
 Operator : MD\SY
 Sample : J5812-24
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TP3-B

Quant Time: Nov 06 06:17:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 22:32:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	331166	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	598342	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	579821	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	255269	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	261884	66.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	133.38%	
35) Dibromofluoromethane	7.59	113	226474	60.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.92%	
50) Toluene-d8	10.09	98	881250	61.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.72%	
62) 4-Bromofluorobenzene	12.40	95	307859	65.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	130.60%	
Target Compounds						
16) Acetone	3.82	43	8424	9.468	ug/l #	87
20) Methylene Chloride	4.56	84	20092	5.435	ug/l	93
43) Isopropyl Acetate	8.17	43	208054	34.449	ug/l #	88
68) m/p-Xylenes	11.62	106	8817	1.164	ug/l	94
95) Naphthalene	15.13	128	194627	24.898	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110518\
Data File : VN052162.D
Acq On : 5 Nov 2018 22:07
Operator : MD\SY
Sample : J5812-24
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TP3-B

Quant Time: Nov 06 06:17:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 22:32:25 2018
Response via : Initial Calibration

