

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110518\
 Data File : VN052166.D
 Acq On : 5 Nov 2018 23:54
 Operator : MD\SY
 Sample : J5815-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 HL-R47183-RT

Manual Integrations
 APPROVED

MMDadoda
 11/6/2018 10:00:46 AM

Quant Time: Nov 06 06:25:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 22:32:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	317677	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	579720	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	578423	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	240201	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	253445	67.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	134.56%	
35) Dibromofluoromethane	7.59	113	220002	61.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.24%	
50) Toluene-d8	10.09	98	863842	62.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.16%	
62) 4-Bromofluorobenzene	12.40	95	295723	64.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	129.48%	
Target Compounds						
16) Acetone	3.82	43	7197m	8.433	ug/l	
20) Methylene Chloride	4.55	84	81434	22.964	ug/l	94
43) Isopropyl Acetate	8.17	43	191001	32.641	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110518\
Data File : VN052166.D
Acq On : 5 Nov 2018 23:54
Operator : MD\SY
Sample : J5815-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
HL-R47183-RT

Manual Integrations
APPROVED

MMDadoda
11/6/2018 10:00:46 AM

Quant Time: Nov 06 06:25:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 22:32:25 2018
Response via : Initial Calibration

