

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110518\
 Data File : VN052170.D
 Acq On : 6 Nov 2018 1:41
 Operator : MD\SY
 Sample : J5805-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-SAMPLES(1-5)

Quant Time: Nov 06 06:35:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 22:32:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	307523	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	565621	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	556034	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	235332	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	247940	68.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	136.00%	
35) Dibromofluoromethane	7.59	113	216505	61.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.30%	
50) Toluene-d8	10.09	98	847155	62.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.78%	
62) 4-Bromofluorobenzene	12.40	95	287887	64.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	129.18%	
Target Compounds						
16) Acetone	3.82	43	122359	148.103	ug/l	97
18) Methyl Acetate	4.34	43	4874	1.856	ug/l #	87
20) Methylene Chloride	4.56	84	23792	6.931	ug/l	94
43) Isopropyl Acetate	8.17	43	164041	28.733	ug/l #	88
95) Naphthalene	15.13	128	7827	1.086	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110518\
Data File : VN052170.D
Acq On : 6 Nov 2018 1:41
Operator : MD\SY
Sample : J5805-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
COMP-SAMPLES(1-5)

Quant Time: Nov 06 06:35:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 22:32:25 2018
Response via : Initial Calibration

