

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110518\
 Data File : VN052172.D
 Acq On : 6 Nov 2018 2:35
 Operator : MD\SY
 Sample : J5805-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-SAMPLES(11-15)

Quant Time: Nov 06 06:39:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 22:32:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	303089	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	556070	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	546102	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	232510	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	241550	67.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	134.42%	
35) Dibromofluoromethane	7.59	113	207901	60.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.44%	
50) Toluene-d8	10.09	98	823739	61.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.42%	
62) 4-Bromofluorobenzene	12.40	95	281625	64.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.54%	
Target Compounds						
16) Acetone	3.81	43	16088	19.758	ug/l	96
20) Methylene Chloride	4.55	84	12244	3.619	ug/l	86
43) Isopropyl Acetate	8.17	43	172709	30.771	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110518\
Data File : VN052172.D
Acq On : 6 Nov 2018 2:35
Operator : MD\SY
Sample : J5805-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
COMP-SAMPLES(11-15)

Quant Time: Nov 06 06:39:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 22:32:25 2018
Response via : Initial Calibration

