

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110518\
 Data File : VN052173.D
 Acq On : 6 Nov 2018 3:01
 Operator : MD\SY
 Sample : J5805-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-SAMPLES(16-20)

Quant Time: Nov 06 06:41:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 22:32:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	308780	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	560013	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	545860	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	236197	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	243059	66.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	132.78%	
35) Dibromofluoromethane	7.59	113	213160	61.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.62%	
50) Toluene-d8	10.09	98	834308	62.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.12%	
62) 4-Bromofluorobenzene	12.40	95	285616	64.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	129.46%	
Target Compounds						
16) Acetone	3.82	43	19336	23.309	ug/l	91
20) Methylene Chloride	4.56	84	23743	6.888	ug/l	92
43) Isopropyl Acetate	8.17	43	171174	30.282	ug/l #	86
68) m/p-Xylenes	11.61	106	9547	1.339	ug/l	98
95) Naphthalene	15.13	128	22707	3.139	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110518\
Data File : VN052173.D
Acq On : 6 Nov 2018 3:01
Operator : MD\SY
Sample : J5805-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
COMP-SAMPLES(16-20)

Quant Time: Nov 06 06:41:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 22:32:25 2018
Response via : Initial Calibration

