

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110519\
 Data File : VN059084.D
 Acq On : 5 Nov 2019 10:49
 Operator : JC/SP
 Sample : VN1105MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1105MBL01

Quant Time: Nov 06 00:47:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 05:43:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	476949	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	740733	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	636556	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	224619	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	303241	48.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.02%	
35) Dibromofluoromethane	7.57	113	234933	49.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.32%	
50) Toluene-d8	10.08	98	908165	48.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.62%	
62) 4-Bromofluorobenzene	12.40	95	280547	41.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN110519\
Data File : VN059084.D
Acq On : 5 Nov 2019 10:49
Operator : JC/SP
Sample : VN1105MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1105MBL01

Quant Time: Nov 06 00:47:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 05:43:23 2019
Response via : Initial Calibration

