

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110519\
 Data File : VN059087.D
 Acq On : 5 Nov 2019 11:55
 Operator : JC/SP
 Sample : K5676-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SU-02-110119

Quant Time: Nov 06 01:13:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110119W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 02 05:43:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	454000	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	712580	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	639631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	230118	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	300691	50.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.02%	
35) Dibromofluoromethane	7.57	113	226824	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
50) Toluene-d8	10.08	98	913954	51.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.14%	
62) 4-Bromofluorobenzene	12.40	95	285209	44.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.66%	
Target Compounds						
16) Acetone	3.80	43	25770	8.466	ug/l	94
20) Methylene Chloride	4.53	84	162415	30.478	ug/l	99
43) Isopropyl Acetate	8.15	43	95920	7.795	ug/l #	92
95) Naphthalene	15.13	128	24541	4.820	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN110519\
Data File : VN059087.D
Acq On : 5 Nov 2019 11:55
Operator : JC/SP
Sample : K5676-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SU-02-110119

Quant Time: Nov 06 01:13:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110119W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 02 05:43:23 2019
Response via : Initial Calibration

