

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110520\
 Data File : VN064460.D
 Acq On : 5 Nov 2020 14:53
 Operator : JC/MD
 Sample : PB132824ZHE#17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132824ZHE#17

Quant Time: Nov 06 03:23:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	247360	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	507440	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	545075	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	199208	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	243061	56.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.82%	
35) Dibromofluoromethane	7.55	113	173586	50.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.48%	
50) Toluene-d8	10.06	98	673884	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	12.38	95	264038	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
Target Compounds						
16) Acetone	3.78	43	12430	7.574	ug/l	99
18) Methyl Acetate	4.30	43	5175	1.604	ug/l #	79
20) Methylene Chloride	4.51	84	8339	2.561	ug/l	97
43) Isopropyl Acetate	8.13	43	105708	12.168	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110520\
Data File : VN064460.D
Acq On : 5 Nov 2020 14:53
Operator : JC/MD
Sample : PB132824ZHE#17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132824ZHE#17

Quant Time: Nov 06 03:23:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

