

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110520\
 Data File : VN064463.D
 Acq On : 5 Nov 2020 16:05
 Operator : JC/MD
 Sample : L4624-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-1

Quant Time: Nov 06 03:29:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	240549	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	496935	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	538315	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	201289	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	230264	54.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.92%	
35) Dibromofluoromethane	7.55	113	167415	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
50) Toluene-d8	10.06	98	661149	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	12.38	95	257110	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	28537	17.880	ug/l	97
18) Methyl Acetate	4.29	43	4966	1.583	ug/l #	87
20) Methylene Chloride	4.51	84	9440	2.981	ug/l	91
40) Benzene	8.01	78	19776	1.282	ug/l	96
43) Isopropyl Acetate	8.13	43	104014	12.226	ug/l	99
52) Toluene	10.13	92	21675	2.304	ug/l	98
95) Naphthalene	15.11	128	33124	6.406	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110520\
Data File : VN064463.D
Acq On : 5 Nov 2020 16:05
Operator : JC/MD
Sample : L4624-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-1

Quant Time: Nov 06 03:29:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

