

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110520\
 Data File : VN064465.D
 Acq On : 5 Nov 2020 16:54
 Operator : JC/MD
 Sample : L4624-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-3

Quant Time: Nov 06 03:33:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	254693	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	534598	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	565052	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	215000	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	244909	55.20	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	110.40%	
35) Dibromofluoromethane	7.55	113	175852	48.31	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	96.62%	
50) Toluene-d8	10.06	98	702485	49.93	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	12.38	95	276534	51.98	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	103.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	26031	15.404	ug/l	98
18) Methyl Acetate	4.29	43	4113	1.238	ug/l	100
20) Methylene Chloride	4.50	84	10105	3.014	ug/l	85
40) Benzene	8.00	78	23352	1.407	ug/l	99
43) Isopropyl Acetate	8.13	43	107547	11.751	ug/l	99
52) Toluene	10.13	92	30379	3.001	ug/l	100
95) Naphthalene	15.11	128	20510	5.236	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110520\
Data File : VN064465.D
Acq On : 5 Nov 2020 16:54
Operator : JC/MD
Sample : L4624-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-3

Quant Time: Nov 06 03:33:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

