

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110520\
 Data File : VN064475.D
 Acq On : 5 Nov 2020 20:57
 Operator : JC/MD
 Sample : L4625-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CONTAINER-002-2A

Quant Time: Nov 06 03:55:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	245831	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	521446	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	555458	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	207285	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	242440	56.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.24%	
35) Dibromofluoromethane	7.55	113	172984	48.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.44%	
50) Toluene-d8	10.06	98	677699	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
62) 4-Bromofluorobenzene	12.38	95	264349	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
Target Compounds						
16) Acetone	3.78	43	23603	14.471	ug/l	95
18) Methyl Acetate	4.30	43	6646	2.073	ug/l #	70
20) Methylene Chloride	4.51	84	9115	2.817	ug/l	99
43) Isopropyl Acetate	8.13	43	102342	11.464	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110520\
Data File : VN064475.D
Acq On : 5 Nov 2020 20:57
Operator : JC/MD
Sample : L4625-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CONTAINER-002-2A

Quant Time: Nov 06 03:55:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

