

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110520\
 Data File : VN064476.D
 Acq On : 5 Nov 2020 21:21
 Operator : JC/MD
 Sample : L4625-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CONTAINER-002-3A

Quant Time: Nov 06 03:57:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	236171	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	495084	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	525658	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	189583	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	231129	56.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.36%	
35) Dibromofluoromethane	7.55	113	167071	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
50) Toluene-d8	10.06	98	642401	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
62) 4-Bromofluorobenzene	12.38	95	251439	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
Target Compounds						
16) Acetone	3.79	43	20366	12.997	ug/l	96
18) Methyl Acetate	4.29	43	5650	1.834	ug/l #	71
20) Methylene Chloride	4.50	84	9150	2.943	ug/l	94
43) Isopropyl Acetate	8.13	43	98363	11.605	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110520\
Data File : VN064476.D
Acq On : 5 Nov 2020 21:21
Operator : JC/MD
Sample : L4625-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CONTAINER-002-3A

Quant Time: Nov 06 03:57:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

