


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN110520\
 Data File : VN064451.D
 Acq On : 5 Nov 2020 10:18
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Title : SW846 8260
 Last Update : Wed Nov 04 12:09:03 2020



AutoFind: Scans 3307, 3308, 3309; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.8	13663	PASS
75	95	30	60	55.9	33520	PASS
95	95	100	100	100.0	60002	PASS
96	95	5	9	6.1	3638	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	70.1	42048	PASS
175	174	5	9	8.2	3467	PASS
176	174	95	101	95.3	40074	PASS
177	176	5	9	6.5	2622	PASS