

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110524\
 Data File : VN084687.D
 Acq On : 05 Nov 2024 14:59
 Operator : JC\MD
 Sample : PB164676TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB164676TB

Quant Time: Nov 06 00:31:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	197070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	344413	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	314716	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	150477	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	135857	47.730	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.460%
35) Dibromofluoromethane	8.165	113	110067	47.214	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.420%
50) Toluene-d8	10.565	98	415024	48.328	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.660%
62) 4-Bromofluorobenzene	12.847	95	162914	50.760	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.520%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	27578	31.043	ug/l	98
20) Methylene Chloride	5.265	84	2568	1.026	ug/l #	76
43) Isopropyl Acetate	8.688	43	118063	21.230	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110524\
Data File : VN084687.D
Acq On : 05 Nov 2024 14:59
Operator : JC\MD
Sample : PB164676TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB164676TB

Quant Time: Nov 06 00:31:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

