

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110524\
 Data File : VN084688.D
 Acq On : 05 Nov 2024 15:23
 Operator : JC\MD
 Sample : P4702-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOTE-1

Quant Time: Nov 06 00:32:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	190206	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	332167	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	292724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	142404	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	131812	47.980	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.960%
35) Dibromofluoromethane	8.159	113	106829	47.514	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.020%
50) Toluene-d8	10.565	98	384997	46.484	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.960%
62) 4-Bromofluorobenzene	12.847	95	154042	49.766	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.540%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	14410	15.483	ug/l	93
30) Chloroform	7.959	83	44024	10.322	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110524\
Data File : VN084688.D
Acq On : 05 Nov 2024 15:23
Operator : JC\MD
Sample : P4702-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOTE-1

Quant Time: Nov 06 00:32:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

