

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110524\
 Data File : VN084690.D
 Acq On : 05 Nov 2024 16:11
 Operator : JC\MD
 Sample : P4693-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-G5-WC

Quant Time: Nov 06 00:33:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	182091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	330511	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	298955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	134248	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	128985	49.044	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.080%
35) Dibromofluoromethane	8.165	113	105459	47.140	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.280%
50) Toluene-d8	10.565	98	390255	47.355	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.720%
62) 4-Bromofluorobenzene	12.847	95	153217	49.747	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.500%
Target Compounds						
16) Acetone	4.430	43	35523	44.413	ug/l	99
43) Isopropyl Acetate	8.688	43	158274	30.049	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110524\
Data File : VN084690.D
Acq On : 05 Nov 2024 16:11
Operator : JC\MD
Sample : P4693-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-G5-WC

Quant Time: Nov 06 00:33:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

