

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110524\
 Data File : VN084691.D
 Acq On : 05 Nov 2024 16:35
 Operator : JC\MD
 Sample : P4693-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-G4-WC

Quant Time: Nov 06 00:34:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	186137	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	333259	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	303530	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	136849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	131974	49.089	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.180%		
35) Dibromofluoromethane	8.165	113	105635	46.829	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.660%		
50) Toluene-d8	10.565	98	395681	47.618	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.240%		
62) 4-Bromofluorobenzene	12.847	95	153000	49.267	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.540%		
Target Compounds						
16) Acetone	4.430	43	34193	41.652	ug/l	98
43) Isopropyl Acetate	8.694	43	181514	34.312	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110524\
Data File : VN084691.D
Acq On : 05 Nov 2024 16:35
Operator : JC\MD
Sample : P4693-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-G4-WC

Quant Time: Nov 06 00:34:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

