

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110524\
 Data File : VN084699.D
 Acq On : 05 Nov 2024 19:48
 Operator : JC\MD
 Sample : P4711-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CF-613-COMP-17

Quant Time: Nov 06 00:38:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	185109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	324473	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	297767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	133611	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	129732	48.523	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.040%		
35) Dibromofluoromethane	8.165	113	103265	47.018	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.040%		
50) Toluene-d8	10.565	98	380097	46.981	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	93.960%		
62) 4-Bromofluorobenzene	12.847	95	149085	49.306	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.620%		
Target Compounds						
16) Acetone	4.430	43	16409	18.607	ug/l	96
25) 2-Butanone	7.494	43	9484	7.604	ug/l	99
43) Isopropyl Acetate	8.706	43	81854	15.364	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110524\
Data File : VN084699.D
Acq On : 05 Nov 2024 19:48
Operator : JC\MD
Sample : P4711-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CF-613-COMP-17

Quant Time: Nov 06 00:38:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

