

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110618\
 Data File : VN052180.D
 Acq On : 6 Nov 2018 9:54
 Operator : MD\SY
 Sample : J5805-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-SAMPLES(26-30)

Quant Time: Nov 06 16:20:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 15:24:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	314009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	566487	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	538327	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	210747	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	241439	64.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	129.68%	
35) Dibromofluoromethane	7.59	113	213864	60.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.62%	
50) Toluene-d8	10.09	98	844025	62.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.14%	
62) 4-Bromofluorobenzene	12.40	95	270906	60.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.38%	
Target Compounds						
20) Methylene Chloride	4.55	84	82123	23.429	ug/l	89
43) Isopropyl Acetate	8.17	43	124808	21.827	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110618\
Data File : VN052180.D
Acq On : 6 Nov 2018 9:54
Operator : MD\SY
Sample : J5805-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
COMP-SAMPLES(26-30)

Quant Time: Nov 06 16:20:05 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 15:24:48 2018
Response via : Initial Calibration

