

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110618\
 Data File : VN052188.D
 Acq On : 6 Nov 2018 13:28
 Operator : MD\SY
 Sample : J5826-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-1

Quant Time: Nov 06 16:36:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 15:24:48 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	311610	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	571375	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	563790	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	238254	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	243333	65.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	131.72%	
35) Dibromofluoromethane	7.59	113	218676	61.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.28%	
50) Toluene-d8	10.09	98	853975	62.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.52%	
62) 4-Bromofluorobenzene	12.40	95	295589	65.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	131.30%	
Target Compounds						
20) Methylene Chloride	4.55	84	12423	3.571	ug/l #	87
43) Isopropyl Acetate	8.17	43	174868	30.321	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110618\
Data File : VN052188.D
Acq On : 6 Nov 2018 13:28
Operator : MD\SY
Sample : J5826-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SB-1

Quant Time: Nov 06 16:36:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 15:24:48 2018
Response via : Initial Calibration

