


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN110618\
Data File : VN052176.D
Acq On : 6 Nov 2018 7:45
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Title : SW846 8260
Last Update : Mon Nov 05 22:32:25 2018



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	13205	PASS
75	95	30	60	53.0	33365	PASS
95	95	100	100	100.0	62930	PASS
96	95	5	9	6.8	4310	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.2	47322	PASS
175	174	5	9	7.8	3668	PASS
176	174	95	101	96.4	45637	PASS
177	176	5	9	6.8	3098	PASS