

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110619\
 Data File : VN059105.D
 Acq On : 6 Nov 2019 14:18
 Operator : JC/SP
 Sample : K5675-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-05-110519

Quant Time: Nov 07 04:38:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	492148	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	763947	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	674102	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	251196	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	303808	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
35) Dibromofluoromethane	7.57	113	234176	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
50) Toluene-d8	10.08	98	948975	52.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.64%	
62) 4-Bromofluorobenzene	12.40	95	303690	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
Target Compounds						
16) Acetone	3.80	43	18856	5.994	ug/l	97
20) Methylene Chloride	4.53	84	445145	81.219	ug/l	99
43) Isopropyl Acetate	8.14	43	105031	8.293	ug/l #	92
95) Naphthalene	15.13	128	6805	3.536	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN110619\
Data File : VN059105.D
Acq On : 6 Nov 2019 14:18
Operator : JC/SP
Sample : K5675-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-05-110519

Quant Time: Nov 07 04:38:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

