

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110619\
 Data File : VN059103.D
 Acq On : 6 Nov 2019 13:34
 Operator : JC/SP
 Sample : VN1106WBSD01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106WBSD01

Manual Integrations
 APPROVED

MMDadoda
 11/7/2019 9:16:32 AM

Quant Time: Nov 07 04:19:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	488524	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	769742	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	687805	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	309511	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	299847	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
35) Dibromofluoromethane	7.57	113	232206	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
50) Toluene-d8	10.08	98	918174	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
62) 4-Bromofluorobenzene	12.40	95	321614	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.83	85	97886	19.887	ug/l	100
3) Chloromethane	2.04	50	130664	19.529	ug/l	98
4) Vinyl Chloride	2.17	62	129143	19.908	ug/l	99
5) Bromomethane	2.55	94	76166	20.404	ug/l	99
6) Chloroethane	2.68	64	76756	20.003	ug/l	99
7) Trichlorofluoromethane	3.00	101	157334	20.278	ug/l	99
8) Diethyl Ether	3.39	74	64774	20.546	ug/l	99
9) 1,1,2-Trichlorotrifluoroet	3.74	101	94875	20.280	ug/l	100
10) Methyl Iodide	3.93	142	116295	20.083	ug/l	99
11) Tert butyl alcohol	4.75	59	118368	103.427	ug/l	99
12) 1,1-Dichloroethene	3.71	96	91821	20.149	ug/l	99
13) Acrolein	3.58	56	59589	100.988	ug/l	97
14) Allyl chloride	4.30	41	187334	20.012	ug/l	98
15) Acrylonitrile	4.96	53	301180	105.745	ug/l	99
16) Acetone	3.79	43	303417	97.161	ug/l	96
17) Carbon Disulfide	4.03	76	277608	19.624	ug/l	100
18) Methyl Acetate	4.30	43	155452	20.972	ug/l	100
19) Methyl tert-butyl Ether	5.01	73	328470	20.935	ug/l	99
20) Methylene Chloride	4.52	84	109734	20.170	ug/l	99
21) trans-1,2-Dichloroethene	5.01	96	99973	20.437	ug/l	97
22) Diisopropyl ether	5.92	45	365792	20.688	ug/l	97
23) Vinyl Acetate	5.87	43	1514870	110.546	ug/l	99
24) 1,1-Dichloroethane	5.82	63	200610	20.314	ug/l	99
25) 2-Butanone	6.81	43	424659	103.180	ug/l	98
26) 2,2-Dichloropropane	6.80	77	167971	19.452	ug/l	99
27) cis-1,2-Dichloroethene	6.81	96	115851	20.414	ug/l	100
28) Bromochloromethane	7.17	49	61415	20.670	ug/l	100
29) Tetrahydrofuran	7.19	42	277764	106.215	ug/l	99
30) Chloroform	7.35	83	192784	20.350	ug/l	99
31) Cyclohexane	7.63	56	185544	19.846	ug/l	97
32) 1,1,1-Trichloroethane	7.55	97	165248	20.747	ug/l	100
36) 1,1-Dichloropropene	7.77	75	145911	20.280	ug/l	99
37) Ethyl Acetate	6.90	43	155344	19.408	ug/l	96
38) Carbon Tetrachloride	7.75	117	142634	19.918	ug/l	99

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110619\
 Data File : VN059103.D
 Acq On : 6 Nov 2019 13:34
 Operator : JC/SP
 Sample : VN1106WBSD01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampled :
 VN1106WBSD01

Manual Integrations
 APPROVED

MMDadoda
 11/7/2019 9:16:32 AM

Quant Time: Nov 07 04:19:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	9.07	83	171124	20.493	ug/l	98
40) Benzene	8.02	78	445238	20.550	ug/l	99
41) Methacrylonitrile	7.15	41	75480	21.687	ug/l #	100
42) 1,2-Dichloroethane	8.10	62	151628	20.405	ug/l	100
43) Isopropyl Acetate	8.14	43	254972	19.981	ug/l	100
44) Trichloroethene	8.82	130	106817	20.056	ug/l	99
45) 1,2-Dichloropropane	9.10	63	120511	20.478	ug/l	98
46) Dibromomethane	9.19	93	73959	20.168	ug/l	99
47) Bromodichloromethane	9.39	83	151669	20.712	ug/l	100
48) Methyl methacrylate	9.18	41	114586	20.076	ug/l	99
49) 1,4-Dioxane	9.18	88	42492	418.910	ug/l	98
51) 4-Methyl-2-Pentanone	9.97	43	813064	112.610	ug/l	99
52) Toluene	10.15	92	264890	20.398	ug/l	100
53) t-1,3-Dichloropropene	10.37	75	168345	19.898	ug/l	97
54) cis-1,3-Dichloropropene	9.83	75	185595	20.394	ug/l	100
55) 1,1,2-Trichloroethane	10.55	97	107904	20.531	ug/l	98
56) Ethyl methacrylate	10.42	69	163569	19.834	ug/l	100
57) 1,3-Dichloropropane	10.70	76	184376	20.383	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.68	63	329149	112.535	ug/l	100
59) 2-Hexanone	10.74	43	604655	109.107	ug/l	100
60) Dibromochloromethane	10.90	129	111990	20.154	ug/l	100
61) 1,2-Dibromoethane	11.00	107	107433	20.193	ug/l	99
64) Tetrachloroethene	10.63	164	96479	20.629	ug/l	96
65) Chlorobenzene	11.43	112	281439	20.990	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.51	131	104419	20.761	ug/l	99
67) Ethyl Benzene	11.51	91	503034	21.146	ug/l	99
68) m/p-Xylenes	11.62	106	366482	41.127	ug/l	98
69) o-Xylene	11.95	106	175545	20.385	ug/l	99
70) Styrene	11.96	104	280642	20.721	ug/l	99
71) Bromoform	12.13	173	78269	20.006	ug/l #	98
73) Isopropylbenzene	12.25	105	478449	21.251	ug/l	100
74) N-amyl acetate	12.07	43	203177	21.297	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.50	83	161605	21.078	ug/l	99
76) 1,2,3-Trichloropropane	12.55	75	145968m	20.715	ug/l	
77) Bromobenzene	12.53	156	113917	20.139	ug/l	98
78) n-propylbenzene	12.59	91	549529	21.317	ug/l	100
79) 2-Chlorotoluene	12.67	91	328583	20.547	ug/l	100
80) 1,3,5-Trimethylbenzene	12.73	105	401373	21.039	ug/l	100
81) trans-1,4-Dichloro-2-buten	12.30	75	54212	20.320	ug/l	97
82) 4-Chlorotoluene	12.77	91	333771	20.810	ug/l	99
83) tert-Butylbenzene	12.99	119	343144	20.829	ug/l	100
84) 1,2,4-Trimethylbenzene	13.04	105	395647	21.436	ug/l	98
85) sec-Butylbenzene	13.17	105	448654	20.940	ug/l	99
86) p-Isopropyltoluene	13.29	119	404246	21.368	ug/l	100
87) 1,3-Dichlorobenzene	13.28	146	199392	20.585	ug/l	98
88) 1,4-Dichlorobenzene	13.36	146	195948	20.219	ug/l	99
89) n-Butylbenzene	13.62	91	341215	20.607	ug/l	99
90) Hexachloroethane	13.88	117	74947	19.729	ug/l	99
91) 1,2-Dichlorobenzene	13.66	146	199604	20.948	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	14.27	75	32571	21.308	ug/l	94

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN110619\
Data File : VN059103.D
Acq On : 6 Nov 2019 13:34
Operator : JC/SP
Sample : VN1106WBSD01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01


Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM

Quant Time: Nov 07 04:19:38 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.91	180	101067	19.464	ug/l	98
94) Hexachlorobutadiene	15.01	225	64222	19.548	ug/l	100
95) Naphthalene	15.13	128	285809	19.780	ug/l	100
96) 1,2,3-Trichlorobenzene	15.30	180	107398	20.112	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1826

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampleId :

VN1106WBSD01

Tgt Ion:168 Resp: 488524

Ion Ratio Lower Upper

168 100


99 56.7 45.7 68.5

Manual Integrations

APPROVED

MMDadoda

11/7/2019 9:16:32 AM



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

7.64

200000



150000

100000

50000

0

Time--> 7.50 7.60 7.70 7.80



#2

Dichlorodifluoromethane

Concen: 19.887 ug/l

RT: 1.83 min Scan# 19

Delta R.T. -0.00 min

Lab File: VN059103.D


Acq: 6 Nov 2019 13:34

Tgt Ion: 85 Resp: 97886

Ion Ratio Lower Upper

85 100

87 32.8 16.3 48.9



Abundance Ion 84.90 (84.60 to 85.60): VN0

Ion 86.90 (86.60 to 87.60): VN0

1.83


60000


40000

20000

0

Time--> 1.75 1.80 1.85 1.90 1.95





#3

Chloromethane

Concen: 19.529 ug/l

RT: 2.04 min Scan# 83

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampleId :

VN1106WBSD01

Tgt Ion: 50 Resp: 130664

Ion Ratio Lower Upper

50 100





52 33.8 26.0 39.0

Manual Integrations

APPROVED

MMDadoda

11/7/2019 9:16:32 AM



#4

Vinyl Chloride

Concen: 19.908 ug/l

RT: 2.17 min Scan# 123

Delta R.T. -0.00 min

Lab File: VN059103.D




Acq: 6 Nov 2019 13:34


Tgt Ion: 62 Resp: 129143

Ion Ratio Lower Upper

62 100

64 31.1 25.4 38.2





#5

Bromomethane

Concen: 20.404 ug/l

RT: 2.55 min Scan# 240

Delta R.T. 0.01 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampleId :

VN1106WBSD01

Tgt Ion: 94 Resp: 76166

Ion Ratio Lower Upper

94 100





96 92.0 74.2 111.2

Manual Integrations

APPROVED

MMDadoda

11/7/2019 9:16:32 AM



#6

Chloroethane

Concen: 20.003 ug/l

RT: 2.68 min Scan# 283

Delta R.T. 0.00 min

Lab File: VN059103.D




Acq: 6 Nov 2019 13:34


Tgt Ion: 64 Resp: 76756

Ion Ratio Lower Upper

64 100

66 30.7 25.2 37.8





#7





Trichlorofluoromethane
 Concen: 20.278 ug/l
 RT: 3.00 min Scan# 380
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106WBSD01

Tgt Ion	Ratio	Lower	Upper
101	100		
103	64.4	51.2	76.8

Manual Integrations
 APPROVED




MMDadoda
 11/7/2019 9:16:32 AM




#8

Diethyl Ether
 Concen: 20.546 ug/l
 RT: 3.39 min Scan# 502
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Lower	Upper
74	100		
45	104.7	51.6	154.9





#9

1,1,2-Trichlorotrifluoroethane

Concen: 20.280 ug/l

RT: 3.74 min Scan# 611

Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N





ClientSampleId :

VN1106WBSD01

Tgt Ion:	101	Resp:	94875
Ion Ratio	Lower	Upper	
101	100		
85	43.9	35.5	53.3
151	77.5	62.2	93.2

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#10

Methyl Iodide

Concen: 20.083 ug/l




RT: 3.93 min Scan# 670


Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion:	142	Resp:	116295
Ion Ratio	Lower	Upper	
142	100		
127	43.0	34.2	51.2
141	13.4	11.4	17.0





#11





Tert butyl alcohol
 Concen: 103.427 ug/l
 RT: 4.75 min Scan# 926
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106WBSD01

Tgt Ion	Ratio	Resp	Lower	Upper
59	100	118368		
57	10.1		8.3	12.5

Manual Integrations
 APPROVED




MMDadoda
 11/7/2019 9:16:32 AM




#12

1,1-Dichloroethene
 Concen: 20.149 ug/l
 RT: 3.71 min Scan# 603
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Resp	Lower	Upper
96	100	91821		
61	171.2		136.5	204.7
98	63.4		50.2	75.4





#13

Acrolein

Concen: 100.988 ug/l

RT: 3.58 min Scan# 563

Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampleId :

VN1106WBSD01

Tgt Ion: 56 Resp: 59589

Ion Ratio Lower Upper

56 100





55 72.9 56.4 84.6

Manual Integrations

APPROVED

MMDadoda

11/7/2019 9:16:32 AM



#14

Allyl chloride

Concen: 20.012 ug/l

RT: 4.30 min Scan# 785

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34




Tgt Ion: 41 Resp: 187334


Ion Ratio Lower Upper

41 100

39 59.7 49.1 73.7

76 31.1 24.9 37.3





#15

Acrylonitrile

Concen: 105.745 ug/l

RT: 4.96 min Scan# 990

Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N





ClientSampleId :

VN1106WBSD01

Tgt Ion:	53	Resp:	301180
Ion	Ratio	Lower	Upper
53	100		
52	82.4	65.0	97.6
51	35.1	28.0	42.0

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#16

Acetone

Concen: 97.161 ug/l




RT: 3.79 min Scan# 627


Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion:	43	Resp:	303417
Ion	Ratio	Lower	Upper
43	100		
58	33.0	24.9	37.3





#17

Carbon Disulfide

Concen: 19.624 ug/l

RT: 4.03 min Scan# 701

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampleId :

VN1106WBSD01

Tgt Ion: 76 Resp: 277608

Ion Ratio Lower Upper

76 100





78 9.1 7.4 11.0

Manual Integrations

APPROVED

MMDadoda

11/7/2019 9:16:32 AM



#18

Methyl Acetate

Concen: 20.972 ug/l

RT: 4.30 min Scan# 785

Delta R.T. -0.00 min

Lab File: VN059103.D




Acq: 6 Nov 2019 13:34


Tgt Ion: 43 Resp: 155452

Ion Ratio Lower Upper

43 100

74 22.4 17.8 26.8





#19

Methyl tert-butyl Ether

Concen: 20.935 ug/l

RT: 5.01 min Scan# 1008

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampleId :

VN1106WBSD01

Tgt Ion: 73 Resp: 328470

Ion Ratio Lower Upper

73 100





57 22.3 18.4 27.6

Manual Integrations

APPROVED

MMDadoda

11/7/2019 9:16:32 AM



#20

Methylene Chloride

Concen: 20.170 ug/l

RT: 4.52 min Scan# 855

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion: 84 Resp: 109734




Ion Ratio Lower Upper


84 100

49 127.8 102.0 153.0

51 37.2 32.2 48.2

86 61.9 50.7 76.1





#21

trans-1,2-Dichloroethene

Concen: 20.437 ug/l

RT: 5.01 min Scan# 1008

Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N





ClientSampleId :

VN1106WBSD01

Tgt Ion:	96	Resp:	99973
Ion	Ratio	Lower	Upper
96	100		
61	150.3	122.6	183.8
98	59.9	50.3	75.5

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#22

Diisopropyl ether

Concen: 20.688 ug/l




RT: 5.92 min Scan# 1291


Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion:	45	Resp:	365792
Ion	Ratio	Lower	Upper
45	100		
43	61.9	46.7	70.1
87	25.3	20.5	30.7
59	11.4	9.6	14.4









#23
 Vinyl Acetate
 Concen: 110.546 ug/l
 RT: 5.87 min Scan# 1273
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Instrument :
 MSVOA_N
 ClientSampleID :
 VN1106WBSD01

Tgt Ion: 43 Resp: 1514870
 Ion Ratio Lower Upper
 43 100
 86 9.3 7.7 11.5




Manual Integrations
 APPROVED


MMDadoda
 11/7/2019 9:16:32 AM



#24
 1,1-Dichloroethane
 Concen: 20.314 ug/l
 RT: 5.82 min Scan# 1259
 Delta R.T. 0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Tgt Ion: 63 Resp: 200610
 Ion Ratio Lower Upper
 63 100
 98 6.1 2.8 8.4
 100 4.0 1.9 5.7









#25
2-Butanone
Concen: 103.180 ug/l
RT: 6.81 min Scan# 1566
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion: 43 Resp: 424659
Ion Ratio Lower Upper
43 100
72 24.6 20.3 30.5




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#26
2,2-Dichloropropane
Concen: 19.452 ug/l
RT: 6.80 min Scan# 1562
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion: 77 Resp: 167971
Ion Ratio Lower Upper
77 100
97 21.7 11.2 33.6





#27

cis-1,2-Dichloroethene

Concen: 20.414 ug/l

RT: 6.81 min Scan# 1565

Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N





ClientSampleId :

VN1106WBSD01

Tgt Ion:	96	Resp:	115851
Ion	Ratio	Lower	Upper
96	100		
61	155.9	0.0	311.6
98	64.1	0.0	129.8

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#28

Bromochloromethane

Concen: 20.670 ug/l




RT: 7.17 min Scan# 1679


Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion:	49	Resp:	61415
Ion	Ratio	Lower	Upper
49	100		
129	1.7	0.0	3.8
130	69.4	55.6	83.4









#29
Tetrahydrofuran
Concen: 106.215 ug/l
RT: 7.19 min Scan# 1684
Delta R.T. 0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion	Ratio	Lower	Upper
42	100		
72	40.5	33.1	49.7
71	38.4	30.9	46.3




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#30
Chloroform
Concen: 20.350 ug/l
RT: 7.35 min Scan# 1734
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Lower	Upper
83	100		
85	65.7	51.8	77.8





#31

Cyclohexane

Concen: 19.846 ug/l

RT: 7.63 min Scan# 1822

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N





ClientSampleId :

VN1106WBSD01

Tgt Ion:	56	Resp:	185544
Ion	Ratio	Lower	Upper
56	100		
69	35.8	25.8	38.8
84	82.2	65.0	97.4

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#32

1,1,1-Trichloroethane

Concen: 20.747 ug/l




RT: 7.55 min Scan# 1796


Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion:	97	Resp:	165248
Ion	Ratio	Lower	Upper
97	100		
99	63.7	51.0	76.6
61	45.1	35.8	53.8









#33
 1,2-Dichloroethane-d4
 Concen: 49.931 ug/l
 RT: 8.00 min Scan# 1938
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106WBSD01

Tgt Ion	Ratio	Resp	Lower	Upper
65	100	299847		
67	54.3	0.0	107.4	




Manual Integrations
 APPROVED


MMDadoda
 11/7/2019 9:16:32 AM



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.57 min Scan# 2113
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Resp	Lower	Upper
114	100	769742		
63	21.8	0.0	43.4	
88	16.3	0.0	32.2	









#35
 Dibromofluoromethane
 Concen: 50.522 ug/l
 RT: 7.57 min Scan# 1802
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106WBSD01

Tgt Ion	Ratio	Resp	Lower	Upper
113	100	232206		
111	102.4	82.8	124.2	
192	20.2	16.1	24.1	




Manual Integrations
 APPROVED


MMDadoda
 11/7/2019 9:16:32 AM



#36
 1,1-Dichloropropene
 Concen: 20.280 ug/l
 RT: 7.77 min Scan# 1866
 Delta R.T. 0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Resp	Lower	Upper
75	100	145911		
110	33.1	17.1	51.2	
77	31.2	24.8	37.2	









#37
Ethyl Acetate
Concen: 19.408 ug/l
RT: 6.90 min Scan# 1595
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampled :
VN1106WBSD01

Tgt Ion	Ratio	Lower	Upper
43	100		
61	15.5	11.2	16.8
70	11.0	7.8	11.8




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#38
Carbon Tetrachloride
Concen: 19.918 ug/l
RT: 7.75 min Scan# 1859
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Lower	Upper
117	100		
119	95.2	75.6	113.4
121	29.7	24.6	36.8









#39
Methylcyclohexane
Concen: 20.493 ug/l
RT: 9.07 min Scan# 2268
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion	Ratio	Lower	Upper
83	100		
55	79.7	63.9	95.9
98	42.0	35.6	53.4




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#40
Benzene
Concen: 20.550 ug/l
RT: 8.02 min Scan# 1943
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Lower	Upper
78	100		
77	23.1	18.8	28.2





#41

Methacrylonitrile

Concen: 21.687 ug/l

RT: 7.15 min Scan# 1671

Delta R.T. -0.01 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampleId :

VN1106WBSD01

Tgt Ion: 41 Resp: 75480

Ion Ratio Lower Upper

41 100





39 60.8 0.0 0.0#

67 81.5 0.0 0.0#

52 31.8 0.0 0.0#

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#42

1,2-Dichloroethane

Concen: 20.405 ug/l

RT: 8.10 min Scan# 1969

Delta R.T. 0.00 min

Lab File: VN059103.D




Acq: 6 Nov 2019 13:34


Tgt Ion: 62 Resp: 151628

Ion Ratio Lower Upper

62 100

98 9.1 0.0 17.8





#43

Isopropyl Acetate

Concen: 19.981 ug/l

RT: 8.14 min Scan# 1981

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N





ClientSampleId :

VN1106WBSD01

Tgt Ion:	43	Resp:	254972
Ion	Ratio	Lower	Upper
43	100		
61	26.3	21.0	31.6
87	12.7	9.8	14.6

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#44

Trichloroethene

Concen: 20.056 ug/l




RT: 8.82 min Scan# 2191


Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion:	130	Resp:	106817
Ion	Ratio	Lower	Upper
130	100		
95	102.2	0.0	206.4









#45
 1,2-Dichloropropane
 Concen: 20.478 ug/l
 RT: 9.10 min Scan# 2279
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106WBSD01

Tgt Ion: 63 Resp: 120511
 Ion Ratio Lower Upper
 63 100
 65 32.0 24.6 37.0




Manual Integrations
 APPROVED


MMDadoda
 11/7/2019 9:16:32 AM



#46
 Dibromomethane
 Concen: 20.168 ug/l
 RT: 9.19 min Scan# 2308
 Delta R.T. 0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Tgt Ion: 93 Resp: 73959
 Ion Ratio Lower Upper
 93 100
 95 83.0 66.6 100.0
 174 98.5 78.2 117.2





#47

Bromodichloromethane

Concen: 20.712 ug/l

RT: 9.39 min Scan# 2368

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N


ClientSampleId :

VN1106WBSD01

Tgt Ion:	83	Resp:	151669
Ion	Ratio	Lower	Upper
83	100		
85	62.1	49.7	74.5
127	7.8	6.4	9.6

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM






Abundance Ion 82.90 (82.60 to 83.60): VN059103.D (-2275) (-)

Ion 84.90 (84.60 to 85.60): VN059103.D (-2275) (-)

Ion 126.80 (126.50 to 127.50): VN059103.D (-2275) (-)

Time-->



#48

Methyl methacrylate

Concen: 20.076 ug/l


RT: 9.18 min Scan# 2305

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion:	41	Resp:	114586
Ion	Ratio	Lower	Upper
41	100		
69	83.4	67.3	100.9
39	53.4	43.3	64.9






Abundance Ion 41.00 (40.70 to 41.70): VN059103.D (-2212) (-)

Ion 69.00 (68.70 to 69.70): VN059103.D (-2212) (-)

Ion 39.00 (38.70 to 39.70): VN059103.D (-2212) (-)

Time-->









#49
1,4-Dioxane
Concen: 418.910 ug/l
RT: 9.18 min Scan# 2305
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion	Ratio	Lower	Upper
88	100		
43	33.8	26.3	39.5
58	74.1	60.6	90.8




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#50
Toluene-d8
Concen: 50.718 ug/l
RT: 10.08 min Scan# 2583
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Lower	Upper
98	100		
100	64.5	52.2	78.2









#51
 4-Methyl-2-Pentanone
 Concen: 112.610 ug/l
 RT: 9.97 min Scan# 2550
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106WBSD01

Tgt Ion: 43 Resp: 813064
 Ion Ratio Lower Upper
 43 100
 58 38.8 31.5 47.3




Manual Integrations
 APPROVED


MMDadoda
 11/7/2019 9:16:32 AM



#52
 Toluene
 Concen: 20.398 ug/l
 RT: 10.15 min Scan# 2604
 Delta R.T. 0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Tgt Ion: 92 Resp: 264890
 Ion Ratio Lower Upper
 92 100
 91 174.5 139.5 209.3





#53

t-1,3-Dichloropropene

Concen: 19.898 ug/l

RT: 10.37 min Scan# 2675

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampleId :

VN1106WBSD01

Tgt Ion: 75 Resp: 168345

Ion Ratio Lower Upper

75 100





77 33.3 25.3 37.9

Manual Integrations

APPROVED

MMDadoda

11/7/2019 9:16:32 AM



#54

cis-1,3-Dichloropropene

Concen: 20.394 ug/l

RT: 9.83 min Scan# 2505

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34




Tgt Ion: 75 Resp: 185595


Ion Ratio Lower Upper

75 100

77 31.7 25.5 38.3

39 46.5 37.3 55.9









#55
 1,1,2-Trichloroethane
 Concen: 20.531 ug/l
 RT: 10.55 min Scan# 2731
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106WBSD01

Tgt Ion	Ratio	Lower	Upper
97	100		
83	86.4	71.0	106.4
85	56.2	45.8	68.8
99	60.8	49.4	74.0




Manual Integrations
 APPROVED


MMDadoda
 11/7/2019 9:16:32 AM



#56
 Ethyl methacrylate
 Concen: 19.834 ug/l
 RT: 10.42 min Scan# 2690
 Delta R.T. 0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Lower	Upper
69	100		
41	66.5	53.6	80.4
39	32.8	26.2	39.4









#57
 1,3-Dichloropropane
 Concen: 20.383 ug/l
 RT: 10.70 min Scan# 2777
 Delta R.T. 0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106WBSD01

Tgt Ion: 76 Resp: 184376
 Ion Ratio Lower Upper
 76 100
 78 32.2 25.8 38.8




Manual Integrations
 APPROVED


MMDadoda
 11/7/2019 9:16:32 AM



#58
 2-Chloroethyl Vinyl ether
 Concen: 112.535 ug/l
 RT: 9.68 min Scan# 2460
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Tgt Ion: 63 Resp: 329149
 Ion Ratio Lower Upper
 63 100
 106 25.0 20.2 30.2









#59
2-Hexanone
Concen: 109.107 ug/l
RT: 10.74 min Scan# 2790
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion: 43 Resp: 604655
Ion Ratio Lower Upper
43 100
58 54.4 27.3 81.8




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#60
Dibromochloromethane
Concen: 20.154 ug/l
RT: 10.90 min Scan# 2837
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion: 129 Resp: 111990
Ion Ratio Lower Upper
129 100
127 77.3 38.6 115.8





#61

1,2-Dibromoethane

Concen: 20.193 ug/l

RT: 11.00 min Scan# 2870

Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampleId :

VN1106WBSD01

Tgt Ion: 107 Resp: 107433

Ion Ratio Lower Upper

107 100


109 95.3 75.5 113.3

Manual Integrations

APPROVED




MMDadoda

11/7/2019 9:16:32 AM



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V



#62

4-Bromofluorobenzene

Concen: 49.271 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34


Tgt Ion: 95 Resp: 321614

Ion Ratio Lower Upper

95 100

174 78.3 0.0 158.2



176 75.5 0.0 151.8




Abundance Ion 94.90 (94.60 to 95.60): V

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.40 min Scan# 2995
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion	Ratio	Resp	Lower	Upper
117	100	687805		
82	58.3	45.8	68.8	
119	31.5	25.8	38.6	




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#64
Tetrachloroethene
Concen: 20.629 ug/l
RT: 10.63 min Scan# 2753
Delta R.T. 0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Resp	Lower	Upper
164	100	96479		
166	123.8	101.5	152.3	
129	93.7	78.3	117.5	
131	89.7	75.9	113.9	









#65
Chlorobenzene
Concen: 20.990 ug/l
RT: 11.43 min Scan# 3003
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion	Ratio	Lower	Upper
112	100		
114	30.8	25.2	37.8




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#66
1,1,1,2-Tetrachloroethane
Concen: 20.761 ug/l
RT: 11.51 min Scan# 3027
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Lower	Upper
131	100		
133	94.4	47.4	142.3
119	67.0	33.2	99.6





#67

Ethyl Benzene

Concen: 21.146 ug/l

RT: 11.51 min Scan# 3028

Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampled :

VN1106WBSD01

Tgt Ion: 91 Resp: 503034





Ion Ratio Lower Upper

91 100

106 30.6 24.1 36.1

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#68

m/p-Xylenes

Concen: 41.127 ug/l

RT: 11.62 min Scan# 3062

Delta R.T. -0.00 min

Lab File: VN059103.D




Acq: 6 Nov 2019 13:34


Tgt Ion: 106 Resp: 366482

Ion Ratio Lower Upper

106 100

91 209.7 164.7 247.1









#69
o-Xylene
Concen: 20.385 ug/l
RT: 11.95 min Scan# 3164
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion:106 Resp: 175545
Ion Ratio Lower Upper
106 100
91 221.5 110.1 330.3




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#70
Styrene
Concen: 20.721 ug/l
RT: 11.96 min Scan# 3169
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion:104 Resp: 280642
Ion Ratio Lower Upper
104 100
78 52.6 41.2 61.8
103 55.4 43.8 65.6









#71
Bromoform
Concen: 20.006 ug/l
RT: 12.13 min Scan# 3220
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion	Ratio	Lower	Upper
173	100		
175	49.7	24.2	72.6
254	0.0	0.1	0.1#




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Lower	Upper
152	100		
115	59.1	29.4	88.2
150	162.2	0.0	345.8









#73
Isopropylbenzene
Concen: 21.251 ug/l
RT: 12.25 min Scan# 3258
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion	Ratio	Lower	Upper
105	100		
120	26.1	13.1	39.3




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#74
N-ethyl acetate
Concen: 21.297 ug/l
RT: 12.07 min Scan# 3202
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Lower	Upper
43	100		
70	41.6	32.2	48.4
55	25.8	20.4	30.6
61	23.6	19.3	28.9





#75

1,1,2,2-Tetrachloroethane

Concen: 21.078 ug/l

RT: 12.50 min Scan# 3337

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampleId :

VN1106WBSD01

Tgt Ion: 83 Resp: 161605
Ion Ratio Lower Upper


83 100

131 9.9 4.9 14.6


85 63.8 32.4 97.2

Manual Integrations
APPROVED



MMDadoda
11/7/2019 9:16:32 AM



Abundance Ion 82.90 (82.60 to 83.60): VN059103.D (-3244) (-)
Ion 130.80 (130.50 to 131.50): VN059103.D (-3244) (-)
Ion 84.90 (84.60 to 85.60): VN059103.D (-3244) (-)



Time-->



#76

1,2,3-Trichloropropane

Concen: 20.715 ug/l m

RT: 12.55 min Scan# 3353

Delta R.T. 0.00 min


Lab File: VN059103.D

Acq: 6 Nov 2019 13:34


Tgt Ion: 75 Resp: 145968
Ion Ratio Lower Upper

75 100


77 0.0 0.0 0.0




Abundance Ion 74.90 (74.60 to 75.60): VN059103.D (-3259) (-)
Ion 77.00 (76.70 to 77.70): VN059103.D (-3259) (-)



Time-->









#77
Bromobenzene
Concen: 20.139 ug/l
RT: 12.53 min Scan# 3344
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion	Ratio	Lower	Upper
156	100		
77	227.6	112.8	338.3
158	98.9	48.4	145.3




Manual Integrations
APPROVED


MMDadoda
11/7/2019 9:16:32 AM



#78
n-propylbenzene
Concen: 21.317 ug/l
RT: 12.59 min Scan# 3364
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Lower	Upper
91	100		
120	22.2	11.2	33.5





#79

2-Chlorotoluene

Concen: 20.547 ug/l

RT: 12.67 min Scan# 3390

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampled :

VN1106WBSD01

Tgt Ion: 91 Resp: 328583

Ion Ratio Lower Upper

91 100





126 33.5 16.7 50.1

Manual Integrations

APPROVED

MMDadoda

11/7/2019 9:16:32 AM



#80

1,3,5-Trimethylbenzene

Concen: 21.039 ug/l

RT: 12.73 min Scan# 3409

Delta R.T. -0.00 min

Lab File: VN059103.D




Acq: 6 Nov 2019 13:34


Tgt Ion: 105 Resp: 401373

Ion Ratio Lower Upper

105 100

120 48.0 24.0 72.0





#81

trans-1,4-Dichloro-2-butene

Concen: 20.320 ug/l

RT: 12.30 min Scan# 3274

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N





ClientSampleId :

VN1106WBSD01

Tgt Ion:	75	Resp:	54212
Ion Ratio	Lower	Upper	
75	100		
53	93.6	72.2	108.2
89	45.5	35.8	53.8

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#82

4-Chlorotoluene

Concen: 20.810 ug/l




RT: 12.77 min Scan# 3421


Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion:	91	Resp:	333771
Ion Ratio	Lower	Upper	
91	100		
126	32.5	16.1	48.2





#83

tert-Butylbenzene

Concen: 20.829 ug/l

RT: 12.99 min Scan# 3490

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N





ClientSampleId :

VN1106WBSD01

Tgt Ion:	119	Resp:	343144
Ion Ratio	Lower	Upper	
119	100		
91	66.6	33.2	99.6
134	25.5	12.7	38.1

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#84

1,2,4-Trimethylbenzene

Concen: 21.436 ug/l




RT: 13.04 min Scan# 3504


Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion:	105	Resp:	395647
Ion Ratio	Lower	Upper	
105	100		
120	44.2	22.8	68.4





#85

sec-Butylbenzene

Concen: 20.940 ug/l

RT: 13.17 min Scan# 3546

Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N

ClientSampleId :

VN1106WBSD01

Tgt Ion:105 Resp: 448654

Ion Ratio Lower Upper

105 100





134 19.8 9.7 29.1

Manual Integrations

APPROVED

MMDadoda

11/7/2019 9:16:32 AM



#86

p-Isopropyltoluene

Concen: 21.368 ug/l

RT: 13.29 min Scan# 3582

Delta R.T. -0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34




Tgt Ion:119 Resp: 404246


Ion Ratio Lower Upper

119 100

134 25.8 13.0 38.9

91 23.6 11.7 35.1









#87
 1,3-Dichlorobenzene
 Concen: 20.585 ug/l
 RT: 13.28 min Scan# 3580
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106WBSD01

Tgt Ion	Ratio	Resp	Lower	Upper
146	100	199392		
111	41.0	20.4	61.3	
148	62.5	32.2	96.6	




Manual Integrations
 APPROVED


MMDadoda
 11/7/2019 9:16:32 AM



#88
 1,4-Dichlorobenzene
 Concen: 20.219 ug/l
 RT: 13.36 min Scan# 3605
 Delta R.T. -0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Resp	Lower	Upper
146	100	195948		
111	40.5	20.3	60.9	
148	64.9	31.9	95.7	





#89

n-Butylbenzene

Concen: 20.607 ug/l

RT: 13.62 min Scan# 3684

Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N





ClientSampleId :

VN1106WBSD01

Tgt Ion:	91	Resp:	341215
Ion	Ratio	Lower	Upper
91	100		
92	51.4	26.1	78.3
134	24.9	12.6	37.8

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#90

Hexachloroethane

Concen: 19.729 ug/l




RT: 13.88 min Scan# 3765


Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion:	117	Resp:	74947
Ion	Ratio	Lower	Upper
117	100		
201	82.8	41.8	125.3





#91

1,2-Dichlorobenzene

Concen: 20.948 ug/l

RT: 13.66 min Scan# 3696

Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Instrument :

MSVOA_N





ClientSampleId :

VN1106WBSD01

Tgt Ion:	146	Resp:	199604
Ion Ratio	Lower	Upper	
146	100		
111	41.6	21.2	63.6
148	63.5	31.6	94.7

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#92

1,2-Dibromo-3-Chloropropane

Concen: 21.308 ug/l




RT: 14.27 min Scan# 3887


Delta R.T. 0.00 min

Lab File: VN059103.D

Acq: 6 Nov 2019 13:34

Tgt Ion:	75	Resp:	32571
Ion Ratio	Lower	Upper	
75	100		
155	76.2	40.2	120.6
157	96.7	51.8	155.5









#93
 1,2,4-Trichlorobenzene
 Concen: 19.464 ug/l
 RT: 14.91 min Scan# 4087
 Delta R.T. 0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106WBSD01

Tgt Ion	Ratio	Resp	Lower	Upper
180	100	101067		
182	97.3	48.1	144.3	
145	31.9	15.1	45.3	




Manual Integrations
 APPROVED


MMDadoda
 11/7/2019 9:16:32 AM



#94
 Hexachlorobutadiene
 Concen: 19.548 ug/l
 RT: 15.01 min Scan# 4118
 Delta R.T. 0.00 min
 Lab File: VN059103.D
 Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Resp	Lower	Upper
225	100	64222		
223	63.0	31.3	93.9	
227	62.8	31.5	94.5	









#95
Naphthalene
Concen: 19.780 ug/l
RT: 15.13 min Scan# 4154
Delta R.T. 0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Instrument :
MSVOA_N
ClientSampleId :
VN1106WBSD01

Tgt Ion	Ratio	Resp	Lower	Upper
128	100	285809		
127	12.8	10.2	15.2	
129	10.8	8.6	12.8	

Manual Integrations
APPROVED

MMDadoda
11/7/2019 9:16:32 AM



#96
1,2,3-Trichlorobenzene
Concen: 20.112 ug/l
RT: 15.30 min Scan# 4206
Delta R.T. -0.00 min
Lab File: VN059103.D
Acq: 6 Nov 2019 13:34

Tgt Ion	Ratio	Resp	Lower	Upper
180	100	107398		
182	95.0	47.3	141.8	
145	31.6	15.8	47.3	

