

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110620\
 Data File : VN064486.D
 Acq On : 6 Nov 2020 16:26
 Operator : JC/MD
 Sample : PB132855ZHE#14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132855ZHE#14

Quant Time: Nov 07 03:48:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	245357	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	505049	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	559026	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	202776	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	240426	56.26	ug/l	0.00
Spiked Amount			Recovery	=	112.52%	
35) Dibromofluoromethane	7.55	113	177617	51.65	ug/l	0.00
Spiked Amount			Recovery	=	103.30%	
50) Toluene-d8	10.06	98	679111	51.10	ug/l	0.00
Spiked Amount			Recovery	=	102.20%	
62) 4-Bromofluorobenzene	12.38	95	265191	52.76	ug/l	0.00
Spiked Amount			Recovery	=	105.52%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Acrolein	3.46	56	66	0.226	ug/l #	1
16) Acetone	3.78	43	11677	7.173	ug/l	93
18) Methyl Acetate	4.30	43	4445	1.389	ug/l #	69
20) Methylene Chloride	4.51	84	10062	3.115	ug/l	93
25) 2-Butanone	6.79	43	717	0.346	ug/l #	52
29) Tetrahydrofuran	7.17	42	440	0.346	ug/l #	48
31) Cyclohexane	7.63	56	6956	1.339	ug/l #	6
36) 1,1-Dichloropropene	7.63	75	24695	4.628	ug/l #	51
37) Ethyl Acetate	6.89	43	220	0.802	ug/l #	71
38) Carbon Tetrachloride	7.63	117	29265	5.132	ug/l #	17
43) Isopropyl Acetate	8.13	43	99807	11.543	ug/l	100
45) 1,2-Dichloropropane	9.24	63	2378	0.568	ug/l #	45
48) Methyl methacrylate	9.14	41	5780	1.325	ug/l #	15
51) 4-Methyl-2-Pentanone	9.96	43	8730	1.731	ug/l	90
54) cis-1,3-Dichloropropene	9.88	75	40750	6.071	ug/l #	67
56) Ethyl methacrylate	10.51	69	3112	2.330	ug/l #	1
57) 1,3-Dichloropropane	10.74	76	6958	1.069	ug/l #	42
58) 2-Chloroethyl Vinyl ether	9.66	63	32	21.548	ug/l #	51
59) 2-Hexanone	10.74	43	120445	32.849	ug/l #	54
74) N-amyl acetate	12.01	43	5775	1.935	ug/l #	52
76) 1,2,3-Trichloropropane	12.38	75	145912	29.699	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.71	105	3066	0.231	ug/l	91
81) trans-1,4-Dichloro-2-buten	12.38	75	145912	82.922	ug/l #	8
95) Naphthalene	15.12	128	342	3.651	ug/l #	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110620\
Data File : VN064486.D
Acq On : 6 Nov 2020 16:26
Operator : JC/MD
Sample : PB132855ZHE#14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132855ZHE#14

Quant Time: Nov 07 03:48:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.63 min Scan# 1831
Delta R.T. 0.00 min
Lab File: VN064486.D
Acq: 6 Nov 2020 16:26

Instrument :
MSVOA_N
ClientSampleId :
PB132855ZHE#14

Tot Ion:168 Resp: 245357
Ion Ratio Lower Upper
168 100
99 70.3 53.4 80.2




Abundance Ion 167.90 (167.60 to 168.60): V
Ion 98.90 (98.60 to 99.60): VN






#13
Acrolein
Concen: 0.226 ug/l
RT: 3.46 min Scan# 536
Delta R.T. -0.10 min
Lab File: VN064486.D
Acq: 6 Nov 2020 16:26

Tgt Ion: 56 Resp: 66
Ion Ratio Lower Upper
56 100
55 166.7 55.9 83.9#



Abundance Ion 56.00 (55.70 to 56.70): VN
Ion 55.00 (54.70 to 55.70): VN









#16
 Acetone
 Concen: 7.173 ug/l
 RT: 3.78 min Scan# 634
 Delta R.T. 0.00 min
 Lab File: VN064486.D
 Acq: 6 Nov 2020 16:26




Instrument :
 MSVOA_N
 ClientSampleId :
 PB132855ZHE#14


Tot Ion: 43 Resp: 11677
 Ion Ratio Lower Upper
 43 100
 58 22.9 21.4 32.0



#18
 Methyl Acetate
 Concen: 1.389 ug/l
 RT: 4.30 min Scan# 795
 Delta R.T. 0.02 min
 Lab File: VN064486.D
 Acq: 6 Nov 2020 16:26

Tgt Ion: 43 Resp: 4445
 Ion Ratio Lower Upper
 43 100
 74 6.9 17.0 25.4#









#20
Methvlene Chloride
Concen: 3.115 ug/l
RT: 4.51 min Scan# 860
Delta R.T. 0.00 min
Lab File: VN064486.D
Acq: 6 Nov 2020 16:26

Instrument :
MSVOA_N
ClientSampleId :
PB132855ZHE#14

Tot Ion:	84	Resp:	10062
Ion	Ratio	Lower	Upper
84	100		
49	126.8	110.1	165.1
51	41.7	33.6	50.4
86	71.4	52.1	78.1




Abundance Ion 83.90 (83.60 to 84.60): VN0
Ion 48.90 (48.60 to 49.60): VN0
Ion 50.90 (50.60 to 51.60): VN0
Ion 85.90 (85.60 to 86.60): VN0






#25
2-Butanone
Concen: 0.346 ug/l
RT: 6.79 min Scan# 1571
Delta R.T. 0.00 min
Lab File: VN064486.D
Acq: 6 Nov 2020 16:26

Tgt Ion:	43	Resp:	717
Ion	Ratio	Lower	Upper
43	100		
72	0.0	18.5	27.7#



Abundance Ion 43.00 (42.70 to 43.70): VN0
Ion 72.00 (71.70 to 72.70): VN0





#29

Tetrahydrofuran

Concen: 0.346 ug/l

RT: 7.17 min Scan# 1690

Delta R.T. 0.01 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26

Instrument :

MSVOA_N

ClientSampleId :

PB132855ZHE#14


Tot Ion: 42 Resp: 440

Ion Ratio Lower Upper

42 100

72 0.0 31.7 47.5#


71 14.1 29.8 44.6#





Abundance Ion 42.00 (41.70 to 42.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

Ion 71.00 (70.70 to 71.70): VN0



Time-->



#31

Cyclohexane

Concen: 1.339 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.02 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26


Tgt Ion: 56 Resp: 6956

Ion Ratio Lower Upper

56 100

69 157.6 25.6 38.4#


84 114.4 63.0 94.6#




Abundance Ion 56.00 (55.70 to 56.70): VN0


Ion 69.00 (68.70 to 69.70): VN0

Ion 84.00 (83.70 to 84.70): VN0



Time-->









#33
 1,2-Dichloroethane-d4
 Concen: 56.257 ug/l
 RT: 7.99 min Scan# 1944
 Delta R.T. 0.00 min
 Lab File: VN064486.D
 Acq: 6 Nov 2020 16:26




Instrument :
 MSVOA_N
 ClientSampleId :
 PB132855ZHE#14


Tot Ion: 65 Resp: 240426
 Ion Ratio Lower Upper
 65 100
 67 50.1 0.0 101.6



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.55 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN064486.D
 Acq: 6 Nov 2020 16:26

Tgt Ion: 114 Resp: 505049
 Ion Ratio Lower Upper
 114 100
 63 22.9 0.0 45.6
 88 16.2 0.0 32.8









#35
Dibromofluoromethane
Concen: 51.653 ug/l
RT: 7.55 min Scan# 1808
Delta R.T. 0.01 min
Lab File: VN064486.D
Acq: 6 Nov 2020 16:26

Instrument :
MSVOA_N
ClientSampleId :
PB132855ZHE#14

Tot Ion:113 Resp: 177617
Ion Ratio Lower Upper
113 100
111 101.2 81.1 121.7
192 16.1 12.8 19.2




Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V






#36
1,1-Dichloropropene
Concen: 4.628 ug/l
RT: 7.63 min Scan# 1831
Delta R.T. -0.13 min
Lab File: VN064486.D
Acq: 6 Nov 2020 16:26

Tgt Ion: 75 Resp: 24695
Ion Ratio Lower Upper
75 100
110 8.5 16.0 48.0#
77 0.0 24.0 36.0#



Abundance Ion 74.90 (74.60 to 75.60): V
Ion 109.90 (109.60 to 110.60): V
Ion 76.90 (76.60 to 77.60): V









#37
Ethyl Acetate
Concen: 0.802 ug/l
RT: 6.89 min Scan# 1603
Delta R.T. 0.01 min
Lab File: VN064486.D
Acq: 6 Nov 2020 16:26




Instrument :
MSVOA_N
ClientSampleId :
PB132855ZHE#14


Tot Ion: 43 Resp: 220
Ion Ratio Lower Upper
43 100
61 0.0 10.0 15.0#
70 0.0 7.1 10.7#



#38
Carbon Tetrachloride
Concen: 5.132 ug/l
RT: 7.63 min Scan# 1833
Delta R.T. -0.10 min
Lab File: VN064486.D
Acq: 6 Nov 2020 16:26

Tgt Ion: 117 Resp: 29265
Ion Ratio Lower Upper
117 100
119 5.5 75.5 113.3#
121 0.0 24.4 36.6#





#43

Isopropyl Acetate

Concen: 11.543 ug/l

RT: 8.13 min Scan# 1987

Delta R.T. 0.00 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26

Instrument :

MSVOA_N

ClientSampleId :

PB132855ZHE#14


Tot Ion: 43 Resp: 99807

Ion Ratio Lower Upper

43 100

61 17.2 14.0 21.0




87 11.4 9.1 13.7



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#45

1,2-Dichloropropane

Concen: 0.568 ug/l

RT: 9.24 min Scan# 2332

Delta R.T. 0.15 min

Lab File: VN064486.D


Acq: 6 Nov 2020 16:26

Tgt Ion: 63 Resp: 2378

Ion Ratio Lower Upper



63 100


65 0.0 23.6 35.4#



Abundance Ion 62.90 (62.60 to 63.60): VN0

Ion 64.90 (64.60 to 65.60): VN0





#48

Methvl methacrylate

Concen: 1.325 ug/l

RT: 9.14 min Scan# 2303

Delta R.T. -0.02 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26

Instrument :

MSVOA_N

ClientSampleId :

PB132855ZHE#14





Tot Ion: 41 Resp: 5780

Ion Ratio Lower Upper

41 100

69 0.0 60.4 90.6#

39 0.0 51.2 76.8#



#50

Toluene-d8

Concen: 51.097 ug/l

RT: 10.06 min Scan# 2588

Delta R.T. 0.00 min

Lab File: VN064486.D




Acq: 6 Nov 2020 16:26


Tgt Ion: 98 Resp: 679111

Ion Ratio Lower Upper

98 100

100 64.3 50.7 76.1





#51

4-Methyl-2-Pentanone

Concen: 1.731 ug/l

RT: 9.96 min Scan# 2555

Delta R.T. 0.00 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26

Instrument :

MSVOA_N

ClientSampleId :





PB132855ZHE#14

Tot Ion: 43 Resp: 8730

Ion Ratio Lower Upper

43 100

58 29.9 28.4 42.6



#54

cis-1,3-Dichloropropene

Concen: 6.071 ug/l

RT: 9.88 min Scan# 2531

Delta R.T. 0.07 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26




Tgt Ion: 75 Resp: 40750


Ion Ratio Lower Upper

75 100

77 0.7 24.9 37.3#

39 42.6 47.1 70.7#





#56

Ethyl methacrylate

Concen: 2.330 ug/l

RT: 10.51 min Scan# 2729

Delta R.T. 0.12 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26

Instrument :

MSVOA_N

ClientSampleId :

PB132855ZHE#14


Tot Ion: 69 Resp: 3112

Ion Ratio Lower Upper

69 100

41 3705.8 59.3 88.9#




39 1693.4 35.4 53.0#



Abundance Ion 69.00 (68.70 to 69.70): VN0

Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0



#57

1,3-Dichloropropane

Concen: 1.069 ug/l

RT: 10.74 min Scan# 2798

Delta R.T. 0.06 min

Lab File: VN064486.D


Acq: 6 Nov 2020 16:26

Tgt Ion: 76 Resp: 6958

Ion Ratio Lower Upper



76 100


78 0.0 25.8 38.6#



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0





#58

2-Chloroethyl Vinyl ether

Concen: 21.548 ug/l

RT: 9.66 min Scan# 2463

Delta R.T. -0.00 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26

Instrument :

MSVOA_N

ClientSampleId :





PB132855ZHE#14

Tot Ion: 63 Resp: 32

Ion Ratio Lower Upper

63 100

106 0.0 19.5 29.3#



#59

2-Hexanone

Concen: 32.849 ug/l

RT: 10.74 min Scan# 2798

Delta R.T. 0.02 min

Lab File: VN064486.D




Acq: 6 Nov 2020 16:26


Tgt Ion: 43 Resp: 120445

Ion Ratio Lower Upper

43 100

58 17.9 24.7 74.1#









#62
4-Bromofluorobenzene
Concen: 52.760 ug/l
RT: 12.38 min Scan# 3308
Delta R.T. 0.00 min
Lab File: VN064486.D
Acq: 6 Nov 2020 16:26




Instrument :
MSVOA_N
ClientSampleId :
PB132855ZHE#14


Tot Ion: 95 Resp: 265191
Ion Ratio Lower Upper
95 100
174 68.9 0.0 132.6
176 65.0 0.0 129.2



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2998
Delta R.T. 0.00 min
Lab File: VN064486.D
Acq: 6 Nov 2020 16:26

Tgt Ion: 117 Resp: 559026
Ion Ratio Lower Upper
117 100
82 60.0 48.1 72.1
119 32.2 26.2 39.4





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3601

Delta R.T. 0.00 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26

Instrument :

MSVOA_N

ClientSampleId :

PB132855ZHE#14


Tot Ion:152 Resp: 202776

Ion Ratio Lower Upper

152 100

115 64.4 32.0 96.0




150 154.7 0.0 353.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-amyl acetate

Concen: 1.935 ug/l

RT: 12.01 min Scan# 3193

Delta R.T. -0.04 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26

Tgt Ion: 43 Resp: 5775


Ion Ratio Lower Upper

43 100

70 0.0 29.6 44.4#

55 9.9 18.4 27.6#

61 0.0 17.8 26.6#






Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V





#76

1,2,3-Trichloropropane

Concen: 29.699 ug/l

RT: 12.38 min Scan# 3308

Delta R.T. -0.15 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26

Instrument :

MSVOA_N

ClientSampleId :





PB132855ZHE#14

Tot Ion: 75 Resp: 145912

Ion Ratio Lower Upper

75 100

77 1.3 0.0 0.0#



#80

1,3,5-Trimethylbenzene

Concen: 0.231 ug/l

RT: 12.71 min Scan# 3411

Delta R.T. 0.00 min

Lab File: VN064486.D




Acq: 6 Nov 2020 16:26


Tgt Ion: 105 Resp: 3066

Ion Ratio Lower Upper

105 100

120 40.5 23.4 70.2





#81

trans-1,4-Dichloro-2-butene

Concen: 82.922 ug/l

RT: 12.38 min Scan# 3308

Delta R.T. 0.10 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26

Instrument :

MSVOA_N

ClientSampleId :

PB132855ZHE#14


Tot Ion: 75 Resp: 145912

Ion Ratio Lower Upper

75 100

53 0.2 84.0 126.0#




89 0.0 36.3 54.5#



Abundance Ion 74.90 (74.60 to 75.60): VN064486.D (-3183) (-)

Ion 53.00 (52.70 to 53.70): VN064486.D (-3183) (-)

Ion 88.90 (88.60 to 89.60): VN064486.D (-3183) (-)



#95

Naphthalene

Concen: 3.651 ug/l

RT: 15.12 min Scan# 4161

Delta R.T. 0.02 min

Lab File: VN064486.D

Acq: 6 Nov 2020 16:26


Tgt Ion: 128 Resp: 342

Ion Ratio Lower Upper

128 100

127 0.0 10.5 15.7#

129 0.0 8.8 13.2#



Abundance Ion 128.00 (127.70 to 128.70): VN064486.D (-4061) (-)

Ion 127.00 (126.70 to 127.70): VN064486.D (-4061) (-)

Ion 129.00 (128.70 to 129.70): VN064486.D (-4061) (-)

