

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110620\
 Data File : VN064493.D
 Acq On : 6 Nov 2020 19:15
 Operator : JC/MD
 Sample : L4656-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SR-L14-L13-NORTH

Quant Time: Nov 07 03:52:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	250273	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	523192	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	558729	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	198916	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	244334	56.05	ug/l	0.00
Spiked Amount			Recovery	=	112.10%	
35) Dibromofluoromethane	7.55	113	178828	50.20	ug/l	0.00
Spiked Amount			Recovery	=	100.40%	
50) Toluene-d8	10.06	98	686270	49.85	ug/l	0.00
Spiked Amount			Recovery	=	99.70%	
62) 4-Bromofluorobenzene	12.38	95	258041	49.56	ug/l	0.00
Spiked Amount			Recovery	=	99.12%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.74	59	122	0.254	ug/l #	71
13) Acrolein	3.57	56	336	1.127	ug/l #	72
16) Acetone	3.78	43	29378	17.692	ug/l	95
18) Methyl Acetate	4.29	43	3199	0.980	ug/l #	89
20) Methylene Chloride	4.51	84	14146	4.294	ug/l	96
25) 2-Butanone	6.80	43	6412	3.031	ug/l	98
29) Tetrahydrofuran	7.17	42	432	0.333	ug/l #	67
31) Cyclohexane	7.63	56	7365	1.390	ug/l #	25
36) 1,1-Dichloropropene	7.63	75	22994	4.160	ug/l #	45
37) Ethyl Acetate	6.94	43	228	0.802	ug/l #	71
38) Carbon Tetrachloride	7.63	117	27972	4.735	ug/l #	19
43) Isopropyl Acetate	8.13	43	93208	10.406	ug/l	99
45) 1,2-Dichloropropane	9.24	63	2070	0.478	ug/l #	45
48) Methyl methacrylate	9.15	41	5725	1.267	ug/l #	15
49) 1,4-Dioxane	9.23	88	99	1.663	ug/l #	1
51) 4-Methyl-2-Pentanone	9.96	43	10871	2.081	ug/l #	76
54) cis-1,3-Dichloropropene	9.87	75	36530	5.254	ug/l #	70
56) Ethyl methacrylate	10.51	69	3127	2.315	ug/l #	1
57) 1,3-Dichloropropane	10.74	76	6257	0.928	ug/l #	42
59) 2-Hexanone	10.74	43	100861	27.325	ug/l #	56
71) Bromoform	12.36	173	431	1.398	ug/l #	1
74) N-amyl acetate	12.00	43	1473	1.402	ug/l #	46
76) 1,2,3-Trichloropropane	12.38	75	144840	30.052	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.38	75	144840	83.910	ug/l #	8
95) Naphthalene	15.10	128	29	3.625	ug/l #	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110620\
Data File : VN064493.D
Acq On : 6 Nov 2020 19:15
Operator : JC/MD
Sample : L4656-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SR-L14-L13-NORTH

Quant Time: Nov 07 03:52:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.00 min

Lab File: VN064493.D

Acq: 6 Nov 2020 19:15

Instrument :

MSVOA_N

ClientSampleId :


SR-L14-L13-NORTH

Tot Ion:168 Resp: 250273

Ion Ratio Lower Upper

168 100

99 67.5 53.4 80.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.63

100000

80000



60000

40000

20000

0

Time--> 7.50 7.60 7.70



#11

Tert butyl alcohol

Concen: 0.254 ug/l

RT: 4.74 min Scan# 934

Delta R.T. -0.00 min

Lab File: VN064493.D


Acq: 6 Nov 2020 19:15

Tgt Ion: 59 Resp: 122

Ion Ratio Lower Upper

59 100

57 0.0 9.0 13.4#



Abundance Ion 59.00 (58.70 to 59.70): VNO

Ion 57.00 (56.70 to 57.70): VNO

4.74

250

200


150


100

50

0

Time--> 4.73 4.74 4.75





#13

Acrolein

Concen: 1.127 ug/l

RT: 3.57 min Scan# 570

Delta R.T. 0.01 min

Lab File: VN064493.D

Acq: 6 Nov 2020 19:15

Instrument :

MSVOA_N

ClientSampleId :





SR-L14-L13-NORTH

Tot Ion: 56 Resp: 336

Ion Ratio Lower Upper

56 100

55 92.9 55.9 83.9#



#16

Acetone

Concen: 17.692 ug/l

RT: 3.78 min Scan# 635

Delta R.T. 0.01 min

Lab File: VN064493.D




Acq: 6 Nov 2020 19:15


Tgt Ion: 43 Resp: 29378

Ion Ratio Lower Upper

43 100

58 24.3 21.4 32.0









#18
Methyl Acetate
Concen: 0.980 ug/l
RT: 4.29 min Scan# 793
Delta R.T. 0.01 min
Lab File: VN064493.D
Acq: 6 Nov 2020 19:15




Instrument :
MSVOA_N
ClientSampleId :
SR-L14-L13-NORTH


Tot Ion: 43 Resp: 3199
Ion Ratio Lower Upper
43 100
74 26.4 17.0 25.4#



#20
Methylene Chloride
Concen: 4.294 ug/l
RT: 4.51 min Scan# 860
Delta R.T. -0.00 min
Lab File: VN064493.D
Acq: 6 Nov 2020 19:15

Tgt Ion: 84 Resp: 14146
Ion Ratio Lower Upper
84 100
49 142.6 110.1 165.1
51 45.8 33.6 50.4
86 62.1 52.1 78.1









#25
 2-Butanone
 Concen: 3.031 ug/l
 RT: 6.80 min Scan# 1573
 Delta R.T. 0.01 min
 Lab File: VN064493.D
 Acq: 6 Nov 2020 19:15




Instrument :
 MSVOA_N
 ClientSampleId :
 SR-L14-L13-NORTH


Tot Ion: 43 Resp: 6412
 Ion Ratio Lower Upper
 43 100
 72 21.9 18.5 27.7



#29
 Tetrahydrofuran
 Concen: 0.333 ug/l
 RT: 7.17 min Scan# 1688
 Delta R.T. -0.00 min
 Lab File: VN064493.D
 Acq: 6 Nov 2020 19:15

Tgt Ion: 42 Resp: 432
 Ion Ratio Lower Upper
 42 100
 72 36.3 31.7 47.5
 71 0.0 29.8 44.6#





#31

Cyclohexane

Concen: 1.390 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. 0.02 min

Lab File: VN064493.D

Acq: 6 Nov 2020 19:15

Instrument :

MSVOA_N

ClientSampleId :

SR-L14-L13-NORTH


Tot Ion: 56 Resp: 7365

Ion Ratio Lower Upper

56 100

69 144.3 25.6 38.4#

84 99.8 63.0 94.6#




Abundance Ion 56.00 (55.70 to 56.70): VN064428.D (-1803) (-)

Ion 69.00 (68.70 to 69.70): VN064428.D (-1803) (-)

Ion 84.00 (83.70 to 84.70): VN064428.D (-1803) (-)

Time-->




Abundance Ion 56.00 (55.70 to 56.70): VN064493.D (-1733) (-)

Ion 69.00 (68.70 to 69.70): VN064493.D (-1733) (-)

Ion 84.00 (83.70 to 84.70): VN064493.D (-1733) (-)

Time-->



#33

1,2-Dichloroethane-d4

Concen: 56.048 ug/l

RT: 7.99 min Scan# 1944

Delta R.T. 0.00 min

Lab File: VN064493.D


Acq: 6 Nov 2020 19:15

Tgt Ion: 65 Resp: 244334

Ion Ratio Lower Upper

65 100


67 49.4 0.0 101.6



Abundance Ion 64.90 (64.60 to 65.60): VN064428.D (-1923) (-)

Ion 66.90 (66.60 to 67.60): VN064428.D (-1923) (-)


Time-->



Abundance Ion 64.90 (64.60 to 65.60): VN064493.D (-1850) (-)

Ion 66.90 (66.60 to 67.60): VN064493.D (-1850) (-)





Time-->



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.55 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN064493.D
 Acq: 6 Nov 2020 19:15




Instrument :
 MSVOA_N
 ClientSampleId :
 SR-L14-L13-NORTH


Tot Ion:114 Resp: 523192
 Ion Ratio Lower Upper
 114 100
 63 23.4 0.0 45.6
 88 16.2 0.0 32.8



#35
 Dibromofluoromethane
 Concen: 50.201 ug/l
 RT: 7.55 min Scan# 1807
 Delta R.T. 0.00 min
 Lab File: VN064493.D
 Acq: 6 Nov 2020 19:15

Tgt Ion:113 Resp: 178828
 Ion Ratio Lower Upper
 113 100
 111 102.1 81.1 121.7
 192 16.0 12.8 19.2









#36
 1,1-Dichloropropene
 Concen: 4.160 ug/l
 RT: 7.63 min Scan# 1831
 Delta R.T. -0.13 min
 Lab File: VN064493.D
 Acq: 6 Nov 2020 19:15




Instrument :
 MSVOA_N
 ClientSampleId :
 SR-L14-L13-NORTH


Tot Ion: 75 Resp: 22994
 Ion Ratio Lower Upper
 75 100
 110 1.6 16.0 48.0#
 77 0.0 24.0 36.0#



#37
 Ethyl Acetate
 Concen: 0.802 ug/l
 RT: 6.94 min Scan# 1616
 Delta R.T. 0.05 min
 Lab File: VN064493.D
 Acq: 6 Nov 2020 19:15

Tgt Ion: 43 Resp: 228
 Ion Ratio Lower Upper
 43 100
 61 0.0 10.0 15.0#
 70 0.0 7.1 10.7#





#38

Carbon Tetrachloride

Concen: 4.735 ug/l

RT: 7.63 min Scan# 1831

Delta R.T. -0.11 min

Lab File: VN064493.D

Acq: 6 Nov 2020 19:15

Instrument :

MSVOA_N

ClientSampleId :

SR-L14-L13-NORTH


Tot Ion:117 Resp: 27972

Ion Ratio Lower Upper

117 100

119 8.0 75.5 113.3#




121 0.0 24.4 36.6#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 10.406 ug/l

RT: 8.13 min Scan# 1987

Delta R.T. 0.00 min

Lab File: VN064493.D

Acq: 6 Nov 2020 19:15


Tgt Ion: 43 Resp: 93208

Ion Ratio Lower Upper

43 100

61 16.7 14.0 21.0



87 11.3 9.1 13.7




Abundance Ion 43.00 (42.70 to 43.70): V

Ion 61.00 (60.70 to 61.70): V

Ion 87.00 (86.70 to 87.70): V





#45

1,2-Dichloropropane

Concen: 0.478 ug/l

RT: 9.24 min Scan# 2333

Delta R.T. 0.16 min

Lab File: VN064493.D

Acq: 6 Nov 2020 19:15

Instrument :

MSVOA_N

ClientSampleId :


SR-L14-L13-NORTH

Tot Ion: 63 Resp: 2070

Ion Ratio Lower Upper




63 100

65 0.0 23.6 35.4#



Abundance Ion 62.90 (62.60 to 63.60): VNO

Ion 64.90 (64.60 to 65.60): VNO



#48

Methyl methacrylate

Concen: 1.267 ug/l

RT: 9.15 min Scan# 2304

Delta R.T. -0.02 min

Lab File: VN064493.D

Acq: 6 Nov 2020 19:15


Tgt Ion: 41 Resp: 5725

Ion Ratio Lower Upper

41 100

69 0.0 60.4 90.6#



39 0.0 51.2 76.8#




Abundance Ion 41.00 (40.70 to 41.70): VNO

Ion 69.00 (68.70 to 69.70): VNO

Ion 39.00 (38.70 to 39.70): VNO





#49

1,4-Dioxane

Concen: 1.663 ug/l

RT: 9.23 min Scan# 2330

Delta R.T. 0.06 min

Lab File: VN064493.D

Acq: 6 Nov 2020 19:15

Instrument :

MSVOA_N

ClientSampleId :

SR-L14-L13-NORTH





Tot Ion: 88 Resp: 99

Ion Ratio Lower Upper

88 100

43 1677414.1 31.0 46.6#

58 8189.9 58.4 87.6#



#50

Toluene-d8

Concen: 49.845 ug/l

RT: 10.06 min Scan# 2588

Delta R.T. 0.00 min

Lab File: VN064493.D




Acq: 6 Nov 2020 19:15


Tgt Ion: 98 Resp: 686270

Ion Ratio Lower Upper

98 100

100 64.0 50.7 76.1





#51

4-Methyl-2-Pentanone

Concen: 2.081 ug/l

RT: 9.96 min Scan# 2556

Delta R.T. 0.01 min

Lab File: VN064493.D

Acq: 6 Nov 2020 19:15

Instrument :

MSVOA_N

ClientSampleId :


SR-L14-L13-NORTH

Tot Ion: 43 Resp: 10871




Ion Ratio Lower Upper

43 100

58 21.4 28.4 42.6#



Abundance Ion 43.00 (42.70 to 43.70): VN064428.D (-2535) (-)



#54

cis-1,3-Dichloropropene

Concen: 5.254 ug/l

RT: 9.87 min Scan# 2530

Delta R.T. 0.07 min

Lab File: VN064493.D

Acq: 6 Nov 2020 19:15


Tgt Ion: 75 Resp: 36530

Ion Ratio Lower Upper



75 100


77 0.0 24.9 37.3#

39 46.8 47.1 70.7#



Abundance Ion 74.90 (74.60 to 75.60): VN064428.D (-2493) (-)





#56

Ethyl methacrylate

Concen: 2.315 ug/l

RT: 10.51 min Scan# 2729

Delta R.T. 0.12 min

Lab File: VN064493.D

Acq: 6 Nov 2020 19:15

Instrument :

MSVOA_N

ClientSampleId :

SR-L14-L13-NORTH





Tot Ion: 69 Resp: 3127

Ion Ratio Lower Upper

69 100

41 3198.8 59.3 88.9#

39 1487.0 35.4 53.0#



#57

1,3-Dichloropropane

Concen: 0.928 ug/l

RT: 10.74 min Scan# 2798

Delta R.T. 0.06 min

Lab File: VN064493.D




Acq: 6 Nov 2020 19:15


Tgt Ion: 76 Resp: 6257

Ion Ratio Lower Upper

76 100

78 0.0 25.8 38.6#









#59
2-Hexanone
Concen: 27.325 ug/l
RT: 10.74 min Scan# 2798
Delta R.T. 0.02 min
Lab File: VN064493.D
Acq: 6 Nov 2020 19:15




Instrument :
MSVOA_N
ClientSampleId :
SR-L14-L13-NORTH


Tot Ion: 43 Resp: 100861
Ion Ratio Lower Upper
43 100
58 19.0 24.7 74.1#



#62
4-Bromofluorobenzene
Concen: 49.558 ug/l
RT: 12.38 min Scan# 3308
Delta R.T. 0.00 min
Lab File: VN064493.D
Acq: 6 Nov 2020 19:15

Tgt Ion: 95 Resp: 258041
Ion Ratio Lower Upper
95 100
174 69.7 0.0 132.6
176 65.5 0.0 129.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2998
Delta R.T. -0.00 min
Lab File: VN064493.D
Acq: 6 Nov 2020 19:15

Instrument :
MSVOA_N
ClientSampleId :
SR-L14-L13-NORTH

Tot Ion:117 Resp: 558729
Ion Ratio Lower Upper
117 100
82 58.2 48.1 72.1
119 31.7 26.2 39.4




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#71
Bromoform
Concen: 1.398 ug/l
RT: 12.36 min Scan# 3304
Delta R.T. 0.26 min
Lab File: VN064493.D
Acq: 6 Nov 2020 19:15

Tgt Ion:173 Resp: 431
Ion Ratio Lower Upper
173 100
175 2627.6 23.7 71.1#
254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V









#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.32 min Scan# 3601
 Delta R.T. 0.00 min
 Lab File: VN064493.D
 Acq: 6 Nov 2020 19:15




Instrument :
 MSVOA_N
 ClientSampleId :
 SR-L14-L13-NORTH


Tot Ion:152 Resp: 198916
 Ion Ratio Lower Upper
 152 100
 115 65.8 32.0 96.0
 150 154.6 0.0 353.8



#74
 N-ethyl acetate
 Concen: 1.402 ug/l
 RT: 12.00 min Scan# 3191
 Delta R.T. -0.04 min
 Lab File: VN064493.D
 Acq: 6 Nov 2020 19:15

Tgt Ion: 43 Resp: 1473
 Ion Ratio Lower Upper
 43 100
 70 0.0 29.6 44.4#
 55 0.0 18.4 27.6#
 61 0.0 17.8 26.6#









#76
 1,2,3-Trichloropropane
 Concen: 30.052 ug/l
 RT: 12.38 min Scan# 3308
 Delta R.T. -0.15 min
 Lab File: VN064493.D
 Acq: 6 Nov 2020 19:15




Instrument :
 MSVOA_N
 ClientSampleId :
 SR-L14-L13-NORTH


Tot Ion: 75 Resp: 144840
 Ion Ratio Lower Upper
 75 100
 77 1.4 0.0 0.0#



#81
 trans-1,4-Dichloro-2-butene
 Concen: 83.910 ug/l
 RT: 12.38 min Scan# 3308
 Delta R.T. 0.10 min
 Lab File: VN064493.D
 Acq: 6 Nov 2020 19:15

Tgt Ion: 75 Resp: 144840
 Ion Ratio Lower Upper
 75 100
 53 0.1 84.0 126.0#
 89 0.0 36.3 54.5#





#95

Naphthalene

Concen: 3.625 ug/l

RT: 15.10 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN064493.D


Acq: 6 Nov 2020 19:15

Instrument :

MSVOA_N

ClientSampleId :

SR-L14-L13-NORTH




Tot Ion:128 Resp: 29

Ion Ratio Lower Upper

128 100

127 0.0 10.5 15.7#


129 0.0 8.8 13.2#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



Time-->