

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110620\
 Data File : VN064484.D
 Acq On : 6 Nov 2020 15:38
 Operator : JC/MD
 Sample : PB132855TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132855TB

Quant Time: Nov 07 04:12:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	236567	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	473387	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	515678	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	182976	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	224615	54.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.02%	
35) Dibromofluoromethane	7.55	113	164546	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
50) Toluene-d8	10.06	98	634682	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	12.38	95	244602	51.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.84%	
Target Compounds						
16) Acetone	3.79	43	17359	11.059	ug/l	92
18) Methyl Acetate	4.29	43	4976	1.613	ug/l #	66
20) Methylene Chloride	4.51	84	16390	5.263	ug/l	97
43) Isopropyl Acetate	8.13	43	95503	11.784	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110620\
Data File : VN064484.D
Acq On : 6 Nov 2020 15:38
Operator : JC/MD
Sample : PB132855TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132855TB

Quant Time: Nov 07 04:12:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

