

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110620\
 Data File : VN064485.D
 Acq On : 6 Nov 2020 16:02
 Operator : JC/MD
 Sample : PB132855ZHE#13
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132855ZHE#13

Quant Time: Nov 07 04:15:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	243377	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	493292	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	553943	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	197420	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	239708	56.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.08%	
35) Dibromofluoromethane	7.55	113	174416	51.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.86%	
50) Toluene-d8	10.06	98	672990	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
62) 4-Bromofluorobenzene	12.38	95	261061	53.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.36%	
Target Compounds						
16) Acetone	3.78	43	13235	8.196	ug/l	95
18) Methyl Acetate	4.30	43	4949m	1.559	ug/l	
20) Methylene Chloride	4.51	84	13799	4.307	ug/l	97
43) Isopropyl Acetate	8.13	43	98676	11.684	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110620\
Data File : VN064485.D
Acq On : 6 Nov 2020 16:02
Operator : JC/MD
Sample : PB132855ZHE#13
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132855ZHE#13

Quant Time: Nov 07 04:15:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

