

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110620\
 Data File : VN064491.D
 Acq On : 6 Nov 2020 18:27
 Operator : JC/MD
 Sample : L4625-22
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CONTAINER-002-4B

Quant Time: Nov 07 04:30:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	236376	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	508403	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	539286	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	194113	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	238450	57.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.82%	
35) Dibromofluoromethane	7.55	113	174445	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
50) Toluene-d8	10.06	98	669537	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	12.38	95	252784	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
Target Compounds						
16) Acetone	3.78	43	19349	12.337	ug/l	97
18) Methyl Acetate	4.29	43	4965	1.610	ug/l #	87
20) Methylene Chloride	4.51	84	9175	2.949	ug/l #	92
43) Isopropyl Acetate	8.13	43	93624	10.756	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110620\
Data File : VN064491.D
Acq On : 6 Nov 2020 18:27
Operator : JC/MD
Sample : L4625-22
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CONTAINER-002-4B

Quant Time: Nov 07 04:30:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

