

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110620\
 Data File : VN064496.D
 Acq On : 6 Nov 2020 20:28
 Operator : JC/MD
 Sample : L4657-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PAV-B49-110520-17-18

Manual Integrations
 APPROVED

MMDadoda
 11/9/2020 9:56:22 AM

Quant Time: Nov 07 04:46:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	243385	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	510341	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	547628	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	200512	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	240790	56.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.60%	
35) Dibromofluoromethane	7.55	113	174081	50.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.20%	
50) Toluene-d8	10.06	98	669170	49.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.66%	
62) 4-Bromofluorobenzene	12.38	95	259133	51.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.04%	
Target Compounds						
16) Acetone	3.78	43	13652	8.454	ug/l	96
18) Methyl Acetate	4.30	43	6303m	1.985	ug/l	
20) Methylene Chloride	4.50	84	13857	4.325	ug/l	97
43) Isopropyl Acetate	8.13	43	98550	11.279	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110620\
Data File : VN064496.D
Acq On : 6 Nov 2020 20:28
Operator : JC/MD
Sample : L4657-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PAV-B49-110520-17-18

Manual Integrations
APPROVED

MMDadoda
11/9/2020 9:56:22 AM

Quant Time: Nov 07 04:46:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

