


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN110620\
 Data File : VN064478.D
 Acq On : 6 Nov 2020 10:29
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Title : SW846 8260
 Last Update : Wed Nov 04 12:09:03 2020



AutoFind: Scans 3307, 3308, 3309; Background Corrected with Scan 3293

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	13849	PASS
75	95	30	60	57.4	35688	PASS
95	95	100	100	100.0	62189	PASS
96	95	5	9	6.2	3825	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	67.5	41973	PASS
175	174	5	9	7.8	3286	PASS
176	174	95	101	96.7	40594	PASS
177	176	5	9	6.5	2635	PASS