

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
 Data File : VN079920.D
 Acq On : 06 Nov 2023 14:10
 Operator : JC\MD
 Sample : PB156885ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156885ZHE#03

Quant Time: Nov 07 04:47:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	244010	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	519797	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	556359	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	202587	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	226234	60.855	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.720%	
35) Dibromofluoromethane	8.171	113	187802	51.048	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.100%	
50) Toluene-d8	10.570	98	698683	51.588	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.180%	
62) 4-Bromofluorobenzene	12.853	95	235703	52.476	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.960%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	102133	78.213	ug/l	91
18) Methyl Acetate	5.024	43	7472	1.196	ug/l #	82
20) Methylene Chloride	5.277	84	22161	6.726	ug/l #	80
25) 2-Butanone	7.488	43	10696	5.594	ug/l #	83
37) Ethyl Acetate	7.571	43	6303	1.308	ug/l #	77
43) Isopropyl Acetate	8.694	43	280248	33.172	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
Data File : VN079920.D
Acq On : 06 Nov 2023 14:10
Operator : JC\MD
Sample : PB156885ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156885ZHE#03

Quant Time: Nov 07 04:47:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

