

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
 Data File : VN079930.D
 Acq On : 06 Nov 2023 18:09
 Operator : JC\MD
 Sample : PB156885ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156885ZHE#13

Quant Time: Nov 07 04:52:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	233548	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	506159	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	548585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	203553	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	223115	62.705	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 125.400%#	
35) Dibromofluoromethane	8.171	113	185955	51.908	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.820%	
50) Toluene-d8	10.571	98	705683	53.509	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.020%	
62) 4-Bromofluorobenzene	12.853	95	235856	53.924	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.840%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	69218	55.381	ug/l	98
18) Methyl Acetate	5.036	43	13837	2.314	ug/l #	74
20) Methylene Chloride	5.277	84	16098	5.104	ug/l #	78
37) Ethyl Acetate	7.571	43	6627	1.413	ug/l #	82
43) Isopropyl Acetate	8.694	43	339953	41.323	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
Data File : VN079930.D
Acq On : 06 Nov 2023 18:09
Operator : JC\MD
Sample : PB156885ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156885ZHE#13

Quant Time: Nov 07 04:52:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

