

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
 Data File : VN079931.D
 Acq On : 06 Nov 2023 18:33
 Operator : JC\MD
 Sample : PB156885ZHE#14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156885ZHE#14

Quant Time: Nov 07 04:52:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	236446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	520173	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	572079	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	210297	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	227779	63.231	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 126.460%#	
35) Dibromofluoromethane	8.171	113	187950	51.052	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.100%	
50) Toluene-d8	10.571	98	710722	52.439	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.880%	
62) 4-Bromofluorobenzene	12.853	95	244688	54.437	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.880%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	59154	46.749	ug/l #	85
18) Methyl Acetate	5.024	43	10844	1.791	ug/l #	59
20) Methylene Chloride	5.289	84	12904	4.042	ug/l	89
37) Ethyl Acetate	7.577	43	6386	1.325	ug/l #	74
43) Isopropyl Acetate	8.694	43	350038	41.402	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
Data File : VN079931.D
Acq On : 06 Nov 2023 18:33
Operator : JC\MD
Sample : PB156885ZHE#14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156885ZHE#14

Quant Time: Nov 07 04:52:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

