

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
 Data File : VN079936.D
 Acq On : 06 Nov 2023 20:33
 Operator : JC\MD
 Sample : PB156885ZHE#19
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156885ZHE#19

Quant Time: Nov 07 04:54:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	234420	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	519988	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	575642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	208024	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	230201	64.456	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 128.920%#			
35) Dibromofluoromethane	8.165	113	191425	52.014	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.020%			
50) Toluene-d8	10.571	98	721428	53.248	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.500%			
62) 4-Bromofluorobenzene	12.853	95	243163	54.116	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 108.240%			
Target Compounds					Qvalue	
16) Acetone	4.436	43	58045	46.269	ug/l	93
18) Methyl Acetate	5.036	43	11133	1.855	ug/l #	74
20) Methylene Chloride	5.289	84	12272	3.877	ug/l #	88
37) Ethyl Acetate	7.565	43	5734	1.190	ug/l #	78
43) Isopropyl Acetate	8.694	43	355808	42.100	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
Data File : VN079936.D
Acq On : 06 Nov 2023 20:33
Operator : JC\MD
Sample : PB156885ZHE#19
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156885ZHE#19

Quant Time: Nov 07 04:54:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

