

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
 Data File : VN079937.D
 Acq On : 06 Nov 2023 20:56
 Operator : JC\MD
 Sample : PB156885ZHE#20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156885ZHE#20

Quant Time: Nov 07 04:55:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	233396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	522851	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	567890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	208549	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	229704	64.599	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 129.200%#	
35) Dibromofluoromethane	8.171	113	189841	51.301	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.600%	
50) Toluene-d8	10.571	98	721710	52.977	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.960%	
62) 4-Bromofluorobenzene	12.853	95	244363	54.086	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.180%	
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	69273	55.461	ug/l	95
18) Methyl Acetate	5.030	43	15300	2.560	ug/l #	81
20) Methylene Chloride	5.277	84	16238	5.152	ug/l #	81
37) Ethyl Acetate	7.571	43	6697	1.382	ug/l #	79
43) Isopropyl Acetate	8.694	43	342609	40.316	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
Data File : VN079937.D
Acq On : 06 Nov 2023 20:56
Operator : JC\MD
Sample : PB156885ZHE#20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156885ZHE#20

Quant Time: Nov 07 04:55:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

