

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
 Data File : VN079946.D
 Acq On : 07 Nov 2023 05:18
 Operator : JC\MD
 Sample : 05256-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-10

Quant Time: Nov 07 06:54:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	271745	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	571071	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	608697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	222001	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	236532	57.132	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.260%
35) Dibromofluoromethane	8.171	113	197341	48.825	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.640%
50) Toluene-d8	10.571	98	732736	49.245	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.500%
62) 4-Bromofluorobenzene	12.853	95	242050	49.050	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.100%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	69881	48.053	ug/l #	88
20) Methylene Chloride	5.283	84	15622	4.257	ug/l	94
37) Ethyl Acetate	7.571	43	5359	1.013	ug/l #	86
43) Isopropyl Acetate	8.694	43	217068	23.386	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
Data File : VN079946.D
Acq On : 07 Nov 2023 05:18
Operator : JC\MD
Sample : 05256-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-10

Quant Time: Nov 07 06:54:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

