

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
 Data File : VN079947.D
 Acq On : 07 Nov 2023 05:42
 Operator : JC\MD
 Sample : 05257-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-6

Quant Time: Nov 07 06:55:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	257640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	549964	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	606968	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	225501	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	229953	58.583	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.160%
35) Dibromofluoromethane	8.171	113	187961	48.289	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.580%
50) Toluene-d8	10.571	98	713245	49.775	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.540%
62) 4-Bromofluorobenzene	12.853	95	242383	51.003	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.000%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	63672	46.180	ug/l	92
18) Methyl Acetate	5.030	43	7488	1.135	ug/l #	67
20) Methylene Chloride	5.277	84	17195	4.942	ug/l #	83
37) Ethyl Acetate	7.571	43	6058	1.189	ug/l #	71
43) Isopropyl Acetate	8.694	43	258700	28.941	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
Data File : VN079947.D
Acq On : 07 Nov 2023 05:42
Operator : JC\MD
Sample : 05257-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-6

Quant Time: Nov 07 06:55:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

