

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
 Data File : VN079948.D
 Acq On : 07 Nov 2023 06:05
 Operator : JC\MD
 Sample : 05257-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2

Quant Time: Nov 07 06:55:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	266751	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	570637	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	610922	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	222487	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	234754	57.764	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.520%
35) Dibromofluoromethane	8.171	113	190116	47.073	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.140%
50) Toluene-d8	10.571	98	720360	48.450	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.900%
62) 4-Bromofluorobenzene	12.853	95	242372	49.153	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.300%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	71132	49.829	ug/l	97
18) Methyl Acetate	5.048	43	8025	1.175	ug/l #	69
20) Methylene Chloride	5.277	84	16682	4.631	ug/l #	79
43) Isopropyl Acetate	8.694	43	228739	24.662	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
Data File : VN079948.D
Acq On : 07 Nov 2023 06:05
Operator : JC\MD
Sample : 05257-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2

Quant Time: Nov 07 06:55:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

