

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
 Data File : VN079949.D
 Acq On : 07 Nov 2023 06:29
 Operator : JC\MD
 Sample : 05257-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-3

Quant Time: Nov 07 06:56:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	271607	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	577915	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	627842	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	233292	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	240403	58.096	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.200%
35) Dibromofluoromethane	8.171	113	196695	48.089	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.180%
50) Toluene-d8	10.571	98	740566	49.182	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.360%
62) 4-Bromofluorobenzene	12.853	95	248961	49.853	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.700%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	65248	44.890	ug/l #	86
18) Methyl Acetate	5.048	43	8733	1.256	ug/l #	72
20) Methylene Chloride	5.277	84	17201	4.690	ug/l #	78
43) Isopropyl Acetate	8.694	43	261093	27.796	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110623\
Data File : VN079949.D
Acq On : 07 Nov 2023 06:29
Operator : JC\MD
Sample : 05257-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-3

Quant Time: Nov 07 06:56:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

