

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110625\
 Data File : VN088217.D
 Acq On : 06 Nov 2025 18:58
 Operator : JC\MD
 Sample : VN1106MBS01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106MBS01

Quant Time: Nov 07 02:25:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	304027	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	601062	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	538359	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	273676	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	317811	60.452	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 120.900%	
35) Dibromofluoromethane	8.147	113	209277	51.827	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.660%	
50) Toluene-d8	10.547	98	775368	49.834	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.660%	
62) 4-Bromofluorobenzene	12.829	95	297482	54.138	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 108.280%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.142	85	61506	17.935	ug/l	97
3) Chloromethane	2.383	50	75414	18.692	ug/l	96
4) Vinyl Chloride	2.536	62	77143	17.549	ug/l	98
5) Bromomethane	2.977	94	44916	17.082	ug/l	88
6) Chloroethane	3.136	64	52383	16.539	ug/l	96
7) Trichlorofluoromethane	3.512	101	116079	16.556	ug/l	95
8) Diethyl Ether	3.965	74	52919	20.308	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.371	101	60126	16.619	ug/l	94
10) Methyl Iodide	4.583	142	58637	17.734	ug/l	95
11) Tert butyl alcohol	5.524	59	99794	110.183	ug/l	96
12) 1,1-Dichloroethene	4.342	96	63273	16.558	ug/l	95
13) Acrolein	4.183	56	71895	70.571	ug/l	98
14) Allyl chloride	5.018	41	127377	19.018	ug/l	97
15) Acrylonitrile	5.712	53	271111	109.890	ug/l	97
16) Acetone	4.424	43	258607	97.914	ug/l	99
17) Carbon Disulfide	4.706	76	198568	17.094	ug/l	95
18) Methyl Acetate	5.018	43	123165	22.778	ug/l	99
19) Methyl tert-butyl Ether	5.789	73	295593	21.063	ug/l	99
20) Methylene Chloride	5.265	84	85837	19.361	ug/l	93
21) trans-1,2-Dichloroethene	5.783	96	77757	18.538	ug/l	95
22) Diisopropyl ether	6.653	45	303187	21.524	ug/l	98
23) Vinyl Acetate	6.588	43	1360915	107.257	ug/l	99
24) 1,1-Dichloroethane	6.553	63	161878	20.746	ug/l	96
25) 2-Butanone	7.465	43	371868	106.838	ug/l	99
26) 2,2-Dichloropropane	7.477	77	129331	18.239	ug/l	99
27) cis-1,2-Dichloroethene	7.471	96	98343	20.366	ug/l	95
28) Bromochloromethane	7.794	49	81690	22.157	ug/l	96
29) Tetrahydrofuran	7.824	42	243631	107.404	ug/l	98
30) Chloroform	7.947	83	162327	20.824	ug/l	100
31) Cyclohexane	8.241	56	127333	17.303	ug/l	96
32) 1,1,1-Trichloroethane	8.153	97	133737	19.263	ug/l	99
36) 1,1-Dichloropropene	8.353	75	106874	18.127	ug/l	99
37) Ethyl Acetate	7.553	43	156709	20.596	ug/l	99
38) Carbon Tetrachloride	8.341	117	109780	18.028	ug/l	99
39) Methylcyclohexane	9.577	83	105721	13.950	ug/l	96
40) Benzene	8.588	78	342444	19.072	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110625\
 Data File : VN088217.D
 Acq On : 06 Nov 2025 18:58
 Operator : JC\MD
 Sample : VN1106MBS01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106MBS01

Quant Time: Nov 07 02:25:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.759	41	76503	19.104	ug/l	97
42) 1,2-Dichloroethane	8.653	62	131794	20.658	ug/l	99
43) Isopropyl Acetate	8.671	43	235543	20.147	ug/l	98
44) Trichloroethene	9.335	130	74863	17.838	ug/l	98
45) 1,2-Dichloropropane	9.600	63	89598	19.765	ug/l	98
46) Dibromomethane	9.694	93	64730	20.022	ug/l	99
47) Bromodichloromethane	9.871	83	130759	20.201	ug/l	98
48) Methyl methacrylate	9.665	41	110974	20.559	ug/l	100
49) 1,4-Dioxane	9.682	88	27691	412.477	ug/l #	98
51) 4-Methyl-2-Pentanone	10.424	43	744568	107.651	ug/l	100
52) Toluene	10.606	92	204738	18.494	ug/l	99
53) t-1,3-Dichloropropene	10.818	75	135418	19.657	ug/l	100
54) cis-1,3-Dichloropropene	10.288	75	145046	19.841	ug/l	98
55) 1,1,2-Trichloroethane	10.994	97	81585	19.425	ug/l	97
56) Ethyl methacrylate	10.853	69	132918	19.998	ug/l	97
57) 1,3-Dichloropropane	11.141	76	147435	19.971	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.141	63	392123	113.421	ug/l	99
59) 2-Hexanone	11.176	43	470368	102.516	ug/l	97
60) Dibromochloromethane	11.335	129	94638	19.825	ug/l	95
61) 1,2-Dibromoethane	11.447	107	81562	18.869	ug/l	96
64) Tetrachloroethene	11.082	164	59673	16.804	ug/l	97
65) Chlorobenzene	11.871	112	225984	18.486	ug/l	93
66) 1,1,1,2-Tetrachloroethane	11.935	131	74967	18.718	ug/l	98
67) Ethyl Benzene	11.941	91	383801	17.886	ug/l	98
68) m/p-Xylenes	12.047	106	283916	35.314	ug/l	99
69) o-Xylene	12.376	106	140003	18.370	ug/l	99
70) Styrene	12.388	104	235213	18.926	ug/l	97
71) Bromoform	12.559	173	55578	18.595	ug/l #	100
73) Isopropylbenzene	12.676	105	347000	16.439	ug/l	99
74) N-amyl acetate	12.547	43	14322	2.390	ug/l #	59
75) 1,1,2,2-Tetrachloroethane	12.918	83	126505	19.250	ug/l	98
76) 1,2,3-Trichloropropane	12.970	75	167993	26.848	ug/l	60
77) Bromobenzene	12.959	156	84655	18.599	ug/l	98
78) n-propylbenzene	13.012	91	415673	16.137	ug/l	100
79) 2-Chlorotoluene	13.100	91	263655	17.948	ug/l	97
80) 1,3,5-Trimethylbenzene	13.153	105	285934	16.311	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.712	75	36048	16.071	ug/l	92
82) 4-Chlorotoluene	13.200	91	274060	17.244	ug/l	99
83) tert-Butylbenzene	13.412	119	242081	15.459	ug/l	98
84) 1,2,4-Trimethylbenzene	13.459	105	296476	16.904	ug/l	100
85) sec-Butylbenzene	13.594	105	333627	14.501	ug/l	99
86) p-Isopropyltoluene	13.706	119	278441	15.086	ug/l	99
87) 1,3-Dichlorobenzene	13.712	146	160278	17.639	ug/l	97
88) 1,4-Dichlorobenzene	13.788	146	162987	16.701	ug/l	96
89) n-Butylbenzene	14.035	91	252713	13.823	ug/l	100
90) Hexachloroethane	14.312	117	52298	14.632	ug/l	97
91) 1,2-Dichlorobenzene	14.082	146	154129	17.597	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.700	75	28153	18.406	ug/l	92
93) 1,2,4-Trichlorobenzene	15.364	180	78547	14.807	ug/l	99
94) Hexachlorobutadiene	15.470	225	24292	12.297	ug/l	98
95) Naphthalene	15.617	128	287798	15.837	ug/l	99
96) 1,2,3-Trichlorobenzene	15.811	180	71791	14.115	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110625\
Data File : VN088217.D
Acq On : 06 Nov 2025 18:58
Operator : JC\MD
Sample : VN1106MBS01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_N/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1106MBS01

Quant Time: Nov 07 02:25:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1063
Delta R.T. 0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBS01


Tgt Ion:168 Resp: 304027
Ion Ratio Lower Upper
168 100
99 70.7 57.2 85.8



#2
Dichlorodifluoromethane
Concen: 17.935 ug/l
RT: 2.142 min Scan# 32
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 85 Resp: 61506
Ion Ratio Lower Upper
85 100
87 31.3 16.6 49.8









#3
Chloromethane
Concen: 18.692 ug/l
RT: 2.383 min Scan# 71
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion: 50 Resp: 75414
Ion Ratio Lower Upper
50 100
52 30.7 26.2 39.4



#4
Vinyl Chloride
Concen: 17.549 ug/l
RT: 2.536 min Scan# 99
Delta R.T. -0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 62 Resp: 77143
Ion Ratio Lower Upper
62 100
64 29.8 24.6 37.0









#5
Bromomethane
Concen: 17.082 ug/l
RT: 2.977 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBS01


Tgt Ion: 94 Resp: 44916
Ion Ratio Lower Upper
94 100
96 82.3 75.4 113.2



#6
Chloroethane
Concen: 16.539 ug/l
RT: 3.136 min Scan# 201
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 64 Resp: 52383
Ion Ratio Lower Upper
64 100
66 30.4 26.2 39.2









#7
Trichlorofluoromethane
Concen: 16.556 ug/l
RT: 3.512 min Scan# 201
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion:101 Resp: 116079
Ion Ratio Lower Upper
101 100
103 65.8 49.4 74.0



#8
Diethyl Ether
Concen: 20.308 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.012 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 74 Resp: 52919
Ion Ratio Lower Upper
74 100
45 100.0 50.6 151.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 16.619 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion:101 Resp: 60126
Ion Ratio Lower Upper
101 100
85 50.1 36.6 55.0
151 72.5 54.7 82.1



#10
Methyl Iodide
Concen: 17.734 ug/l
RT: 4.583 min Scan# 447
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion:142 Resp: 58637
Ion Ratio Lower Upper
142 100
127 51.3 38.0 57.0
141 16.2 12.0 18.0









#11
Tert butyl alcohol
Concen: 110.183 ug/l
RT: 5.524 min Scan# 607
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion: 59 Resp: 99794
Ion Ratio Lower Upper
59 100
57 9.7 9.0 13.4



#12
1,1-Dichloroethene
Concen: 16.558 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 96 Resp: 63273
Ion Ratio Lower Upper
96 100
61 188.3 143.8 215.8
98 63.4 52.1 78.1









#13
Acrolein
Concen: 70.571 ug/l
RT: 4.183 min Scan# 31
Delta R.T. 0.012 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion: 56 Resp: 71895
Ion Ratio Lower Upper
56 100
55 73.5 57.3 85.9



#14
Allyl chloride
Concen: 19.018 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.012 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 41 Resp: 127377
Ion Ratio Lower Upper
41 100
39 82.4 63.9 95.9
76 36.1 28.1 42.1









#15
Acrylonitrile
Concen: 109.890 ug/l
RT: 5.712 min Scan# 611
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion: 53 Resp: 271111
Ion Ratio Lower Upper
53 100
52 82.0 67.6 101.4
51 36.0 30.1 45.1



#16
Acetone
Concen: 97.914 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 43 Resp: 258607
Ion Ratio Lower Upper
43 100
58 27.9 22.6 33.8









#17
Carbon Disulfide
Concen: 17.094 ug/l
RT: 4.706 min Scan# 41
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion: 76 Resp: 198568
Ion Ratio Lower Upper
76 100
78 7.7 7.6 11.4



#18
Methyl Acetate
Concen: 22.778 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 43 Resp: 123165
Ion Ratio Lower Upper
43 100
74 22.4 17.6 26.4









#19
Methyl tert-butyl Ether
Concen: 21.063 ug/l
RT: 5.789 min Scan# 611
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBS01


Tgt Ion: 73 Resp: 295593
Ion Ratio Lower Upper
73 100
57 22.5 18.4 27.6



#20
Methylene Chloride
Concen: 19.361 ug/l
RT: 5.265 min Scan# 563
Delta R.T. 0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 84 Resp: 85837
Ion Ratio Lower Upper
84 100
49 140.1 103.4 155.2
51 40.9 33.8 50.8
86 66.9 49.6 74.4









#21
trans-1,2-Dichloroethene
Concen: 18.538 ug/l
RT: 5.783 min Scan# 611
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion: 96 Resp: 77757
Ion Ratio Lower Upper
96 100
61 141.5 118.0 177.0
98 57.5 50.0 75.0



#22
Diisopropyl ether
Concen: 21.524 ug/l
RT: 6.653 min Scan# 799
Delta R.T. -0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 45 Resp: 303187
Ion Ratio Lower Upper
45 100
43 60.9 47.7 71.5
87 27.7 21.4 32.0
59 12.2 8.7 13.1









#23
 Vinyl Acetate
 Concen: 107.257 ug/l
 RT: 6.588 min Scan# 788
 Delta R.T. -0.000 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106MBS01

Tgt Ion: 43 Resp: 1360915
 Ion Ratio Lower Upper
 43 100
 86 10.0 8.2 12.4



#24
 1,1-Dichloroethane
 Concen: 20.746 ug/l
 RT: 6.553 min Scan# 782
 Delta R.T. 0.006 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58

Tgt Ion: 63 Resp: 161878
 Ion Ratio Lower Upper
 63 100
 98 5.7 3.8 11.4
 100 4.4 2.0 6.0









#25
2-Butanone
Concen: 106.838 ug/l
RT: 7.465 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBS01

Tgt Ion: 43 Resp: 371868
Ion Ratio Lower Upper
43 100
72 24.2 19.9 29.9



#26
2,2-Dichloropropane
Concen: 18.239 ug/l
RT: 7.477 min Scan# 939
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 77 Resp: 129331
Ion Ratio Lower Upper
77 100
97 21.2 10.9 32.7









#27
 cis-1,2-Dichloroethene
 Concen: 20.366 ug/l
 RT: 7.471 min Scan# 91
 Delta R.T. 0.006 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106MBS01

Tgt Ion: 96 Resp: 98343




Ion	Ratio	Lower	Upper
96	100		
61	159.3	0.0	301.8
98	64.1	0.0	126.2




#28
 Bromochloromethane
 Concen: 22.157 ug/l
 RT: 7.794 min Scan# 993
 Delta R.T. -0.006 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58

Tgt Ion: 49 Resp: 81690

Ion	Ratio	Lower	Upper
49	100		
129	0.2	0.0	3.0
130	62.9	53.1	79.7





#29

Tetrahydrofuran

Concen: 107.404 ug/l

RT: 7.824 min Scan# 998

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :

VN1106MBS01





Tgt Ion: 42 Resp: 243631

Ion Ratio Lower Upper

42 100

72 41.0 33.7 50.5

71 38.1 31.9 47.9



#30

Chloroform

Concen: 20.824 ug/l

RT: 7.947 min Scan# 1019

Delta R.T. -0.000 min

Lab File: VN088217.D




Acq: 06 Nov 2025 18:58


Tgt Ion: 83 Resp: 162327

Ion Ratio Lower Upper

83 100

85 63.2 50.7 76.1









#31
Cyclohexane
Concen: 17.303 ug/l
RT: 8.241 min Scan# 1069
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBS01


Tgt Ion: 56 Resp: 127333
Ion Ratio Lower Upper
56 100
69 28.4 23.7 35.5
84 77.3 64.7 97.1



#32
1,1,1-Trichloroethane
Concen: 19.263 ug/l
RT: 8.153 min Scan# 1054
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 97 Resp: 133737
Ion Ratio Lower Upper
97 100
99 65.6 52.2 78.2
61 50.0 41.2 61.8





#33

1,2-Dichloroethane-d4

Concen: 60.452 ug/l

RT: 8.559 min Scan# 1123

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :





VN1106MBS01

Tgt Ion: 65 Resp: 317811

Ion Ratio Lower Upper

65 100

67 51.0 0.0 100.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.082 min Scan# 1212

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58




Tgt Ion: 114 Resp: 601062


Ion Ratio Lower Upper

114 100

63 22.6 0.0 45.2

88 16.9 0.0 33.4









#35
Dibromofluoromethane
Concen: 51.827 ug/l
RT: 8.147 min Scan# 1053
Delta R.T. -0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBS01


Tgt Ion:113 Resp: 209277
Ion Ratio Lower Upper
113 100
111 105.7 82.8 124.2
192 16.2 12.8 19.2



#36
1,1-Dichloropropene
Concen: 18.127 ug/l
RT: 8.353 min Scan# 1088
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 75 Resp: 106874
Ion Ratio Lower Upper
75 100
110 32.6 15.9 47.6
77 31.2 24.4 36.6





#37

Ethyl Acetate

Concen: 20.596 ug/l

RT: 7.553 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :

VN1106MBS01





Tgt Ion: 43 Resp: 156709

Ion Ratio Lower Upper

43 100

61 11.6 9.3 13.9

70 9.6 7.0 10.6



#38

Carbon Tetrachloride

Concen: 18.028 ug/l

RT: 8.341 min Scan# 1086

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58




Tgt Ion: 117 Resp: 109780


Ion Ratio Lower Upper

117 100

119 99.7 79.0 118.6

121 30.6 25.0 37.6





#39

Methylcyclohexane

Concen: 13.950 ug/l

RT: 9.577 min Scan# 1129

Delta R.T. -0.006 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :

VN1106MBSD01





Tgt Ion: 83 Resp: 105721

Ion Ratio Lower Upper

83 100

55 86.1 64.6 96.8

98 48.8 38.4 57.6



#40

Benzene

Concen: 19.072 ug/l

RT: 8.588 min Scan# 1128

Delta R.T. -0.000 min

Lab File: VN088217.D




Acq: 06 Nov 2025 18:58

Tgt Ion: 78 Resp: 342444

Ion Ratio Lower Upper

78 100


77 24.5 19.7 29.5








#41
Methacrylonitrile
Concen: 19.104 ug/l
RT: 7.759 min Scan# 91
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01







Tgt Ion:	41	Resp:	76503
Ion	Ratio	Lower	Upper
41	100		
39	56.9	45.4	68.0
67	72.3	55.3	82.9
52	33.8	24.5	36.7



#42
1,2-Dichloroethane
Concen: 20.658 ug/l
RT: 8.653 min Scan# 1139
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion:	62	Resp:	131794
Ion	Ratio	Lower	Upper
62	100		
98	8.4	0.0	17.4









#43
Isopropyl Acetate
Concen: 20.147 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion: 43 Resp: 235543
Ion Ratio Lower Upper
43 100
61 25.2 18.8 28.2
87 12.1 9.6 14.4



#44
Trichloroethene
Concen: 17.838 ug/l
RT: 9.335 min Scan# 1255
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion:130 Resp: 74863
Ion Ratio Lower Upper
130 100
95 108.3 0.0 212.0





#45

1,2-Dichloropropane

Concen: 19.765 ug/l

RT: 9.600 min Scan# 1300

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :





VN1106MBS01

Tgt Ion: 63 Resp: 89598

Ion Ratio Lower Upper

63 100

65 30.3 25.0 37.4



#46

Dibromomethane

Concen: 20.022 ug/l

RT: 9.694 min Scan# 1316

Delta R.T. 0.006 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58




Tgt Ion: 93 Resp: 64730


Ion Ratio Lower Upper

93 100

95 82.4 66.2 99.4

174 80.6 63.4 95.2





#47

Bromodichloromethane

Concen: 20.201 ug/l

RT: 9.871 min Scan# 1346

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :

VN1106MBSD01





Tgt Ion: 83 Resp: 130759

Ion Ratio Lower Upper

83 100

85 63.8 52.5 78.7

127 8.1 5.8 8.6



#48

Methyl methacrylate

Concen: 20.559 ug/l

RT: 9.665 min Scan# 1311

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58




Tgt Ion: 41 Resp: 110974


Ion Ratio Lower Upper

41 100

69 79.0 62.7 94.1

39 65.9 52.9 79.3





#49

1,4-Dioxane

Concen: 412.477 ug/l

RT: 9.682 min Scan# 1314

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :

VN1106MBSD01





Tgt Ion: 88 Resp: 27691

Ion Ratio Lower Upper

88 100

43 23.1 0.0 0.0#

58 74.4 57.9 86.9



#50

Toluene-d8

Concen: 49.834 ug/l

RT: 10.547 min Scan# 1461

Delta R.T. -0.000 min

Lab File: VN088217.D




Acq: 06 Nov 2025 18:58


Tgt Ion: 98 Resp: 775368

Ion Ratio Lower Upper

98 100

100 64.6 52.8 79.2









#51
4-Methyl-2-Pentanone
Concen: 107.651 ug/l
RT: 10.424 min Scan# 1440
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion: 43 Resp: 744568
Ion Ratio Lower Upper
43 100
58 36.8 29.6 44.4



#52
Toluene
Concen: 18.494 ug/l
RT: 10.606 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 92 Resp: 204738
Ion Ratio Lower Upper
92 100
91 173.6 140.4 210.6









#53
t-1,3-Dichloropropene
Concen: 19.657 ug/l
RT: 10.818 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion: 75 Resp: 135418
Ion Ratio Lower Upper
75 100
77 31.3 25.2 37.8



#54
cis-1,3-Dichloropropene
Concen: 19.841 ug/l
RT: 10.288 min Scan# 1417
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 75 Resp: 145046
Ion Ratio Lower Upper
75 100
77 30.2 24.6 37.0
39 55.0 45.2 67.8





#55

1,1,2-Trichloroethane

Concen: 19.425 ug/l

RT: 10.994 min Scan# 1537

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :

VN1106MBS01

Tgt Ion: 97 Resp: 81585





Ion Ratio Lower Upper

97 100

83 85.0 68.2 102.4

85 59.2 45.5 68.3

99 60.5 52.2 78.4



#56

Ethyl methacrylate

Concen: 19.998 ug/l

RT: 10.853 min Scan# 1513

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58




Tgt Ion: 69 Resp: 132918


Ion Ratio Lower Upper

69 100

41 71.8 56.5 84.7

39 45.5 34.2 51.4





#57

1,3-Dichloropropane

Concen: 19.971 ug/l

RT: 11.141 min Scan# 1562

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :





VN1106MBSD01

Tgt Ion: 76 Resp: 147435

Ion Ratio Lower Upper

76 100

78 32.0 26.2 39.4



#58

2-Chloroethyl Vinyl ether

Concen: 113.421 ug/l

RT: 10.141 min Scan# 1392

Delta R.T. -0.000 min

Lab File: VN088217.D




Acq: 06 Nov 2025 18:58


Tgt Ion: 63 Resp: 392123

Ion Ratio Lower Upper

63 100

106 27.0 21.1 31.7









#59
2-Hexanone
Concen: 102.516 ug/l
RT: 11.176 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion: 43 Resp: 470368
Ion Ratio Lower Upper
43 100
58 53.7 25.9 77.8



#60
Dibromochloromethane
Concen: 19.825 ug/l
RT: 11.335 min Scan# 1595
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 129 Resp: 94638
Ion Ratio Lower Upper
129 100
127 73.5 39.1 117.3









#61
1,2-Dibromoethane
Concen: 18.869 ug/l
RT: 11.447 min Scan# 1614
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion:107 Resp: 81562
Ion Ratio Lower Upper
107 100
109 96.9 74.7 112.1



#62
4-Bromofluorobenzene
Concen: 54.138 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 95 Resp: 297482
Ion Ratio Lower Upper
95 100
174 66.9 0.0 138.4
176 63.9 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.841 min Scan# 1170
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion:117 Resp: 538359
Ion Ratio Lower Upper
117 100
82 59.7 47.4 71.2
119 31.6 24.3 36.5



#64
Tetrachloroethene
Concen: 16.804 ug/l
RT: 11.082 min Scan# 1552
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion:164 Resp: 59673
Ion Ratio Lower Upper
164 100
166 121.5 97.7 146.5
129 106.2 85.9 128.9
131 97.2 83.2 124.8









#65
Chlorobenzene
Concen: 18.486 ug/l
RT: 11.871 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01

Tgt Ion:112 Resp: 225984
Ion Ratio Lower Upper
112 100
114 34.4 24.3 36.5



#66
1,1,1,2-Tetrachloroethane
Concen: 18.718 ug/l
RT: 11.935 min Scan# 1697
Delta R.T. -0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion:131 Resp: 74967
Ion Ratio Lower Upper
131 100
133 97.4 48.0 144.0
119 67.6 32.6 97.7









#67
Ethyl Benzene
Concen: 17.886 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBS01


Tgt Ion: 91 Resp: 383801
Ion Ratio Lower Upper
91 100
106 29.2 24.1 36.1



#68
m/p-Xylenes
Concen: 35.314 ug/l
RT: 12.047 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 106 Resp: 283916
Ion Ratio Lower Upper
106 100
91 212.8 169.3 253.9









#69
o-Xylene
Concen: 18.370 ug/l
RT: 12.376 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion:106 Resp: 140003
Ion Ratio Lower Upper
106 100
91 221.6 111.4 334.2



#70
Styrene
Concen: 18.926 ug/l
RT: 12.388 min Scan# 1774
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion:104 Resp: 235213
Ion Ratio Lower Upper
104 100
78 55.8 42.2 63.2
103 55.0 43.4 65.2









#71
Bromoform
Concen: 18.595 ug/l
RT: 12.559 min Scan# 1803
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion:173 Resp: 55578
Ion Ratio Lower Upper
173 100
175 49.9 24.9 74.6
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion:152 Resp: 273676
Ion Ratio Lower Upper
152 100
115 65.7 32.3 96.8
150 162.1 0.0 351.0





#73

Isopropylbenzene

Concen: 16.439 ug/l

RT: 12.676 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :





VN1106MBSD01

Tgt Ion:105 Resp: 347000

Ion Ratio Lower Upper

105 100

120 24.8 12.8 38.3



#74

N-ethyl acetate

Concen: 2.390 ug/l

RT: 12.547 min Scan# 1801

Delta R.T. 0.047 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Tgt Ion: 43 Resp: 14322




Ion Ratio Lower Upper


43 100

70 0.0 27.6 41.4#

55 30.3 13.0 19.6#

61 11.8 13.0 19.6#









#75
 1,1,2-Tetrachloroethane
 Concen: 19.250 ug/l
 RT: 12.918 min Scan# 1864
 Delta R.T. -0.000 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106MBS01


Tgt Ion: 83 Resp: 126505
 Ion Ratio Lower Upper
 83 100
 131 9.4 4.8 14.4
 85 62.4 32.1 96.3



#76
 1,2,3-Trichloropropane
 Concen: 26.848 ug/l
 RT: 12.970 min Scan# 1873
 Delta R.T. -0.000 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58

Tgt Ion: 75 Resp: 167993
 Ion Ratio Lower Upper
 75 100
 77 122.6 89.1 267.4





#77

Bromobenzene

Concen: 18.599 ug/l

RT: 12.959 min Scan# 1871

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :

VN1106MBS01





Tgt Ion:156 Resp: 84655

Ion Ratio Lower Upper

156 100

77 243.3 119.3 358.0

158 97.7 48.2 144.6



#78

n-propylbenzene

Concen: 16.137 ug/l

RT: 13.012 min Scan# 1880

Delta R.T. -0.000 min

Lab File: VN088217.D




Acq: 06 Nov 2025 18:58


Tgt Ion: 91 Resp: 415673

Ion Ratio Lower Upper

91 100

120 21.4 10.7 32.1









#79
2-Chlorotoluene
Concen: 17.948 ug/l
RT: 13.100 min Scan# 1895
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBS01


Tgt Ion: 91 Resp: 263655
Ion Ratio Lower Upper
91 100
126 30.7 16.1 48.3



#80
1,3,5-Trimethylbenzene
Concen: 16.311 ug/l
RT: 13.153 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 105 Resp: 285934
Ion Ratio Lower Upper
105 100
120 47.8 23.2 69.6









#81
trans-1,4-Dichloro-2-butene
Concen: 16.071 ug/l
RT: 12.712 min Scan# 1829
Delta R.T. -0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBS01


Tgt Ion: 75 Resp: 36048
Ion Ratio Lower Upper
75 100
53 116.5 101.8 152.6
89 47.6 35.6 53.4



#82
4-Chlorotoluene
Concen: 17.244 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion: 91 Resp: 274060
Ion Ratio Lower Upper
91 100
126 31.7 15.6 46.8





#83

tert-Butylbenzene

Concen: 15.459 ug/l

RT: 13.412 min Scan# 1948

Delta R.T. -0.006 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :

VN1106MBS01





Tgt Ion:119 Resp: 242081

Ion Ratio Lower Upper

119 100

91 68.7 35.0 105.0

134 26.6 12.6 37.8



#84

1,2,4-Trimethylbenzene

Concen: 16.904 ug/l

RT: 13.459 min Scan# 1956

Delta R.T. -0.000 min

Lab File: VN088217.D




Acq: 06 Nov 2025 18:58


Tgt Ion:105 Resp: 296476

Ion Ratio Lower Upper

105 100

120 43.8 22.0 66.0









#85
 sec-Butylbenzene
 Concen: 14.501 ug/l
 RT: 13.594 min Scan# 1979
 Delta R.T. -0.000 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106MBSD01


Tgt Ion:105 Resp: 333627
 Ion Ratio Lower Upper
 105 100
 134 18.8 9.6 28.6



#86
 p-Isopropyltoluene
 Concen: 15.086 ug/l
 RT: 13.706 min Scan# 1998
 Delta R.T. -0.000 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58

Tgt Ion:119 Resp: 278441
 Ion Ratio Lower Upper
 119 100
 134 25.6 12.7 38.0
 91 26.2 12.5 37.5









#87
 1,3-Dichlorobenzene
 Concen: 17.639 ug/l
 RT: 13.712 min Scan# 1999
 Delta R.T. -0.000 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106MBSD01


Tgt Ion:146 Resp: 160278
 Ion Ratio Lower Upper
 146 100
 111 43.7 22.1 66.5
 148 61.1 32.0 96.2



#88
 1,4-Dichlorobenzene
 Concen: 16.701 ug/l
 RT: 13.788 min Scan# 1212
 Delta R.T. -0.000 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58

Tgt Ion:146 Resp: 162987
 Ion Ratio Lower Upper
 146 100
 111 45.4 21.0 63.0
 148 66.2 32.0 96.0





#89

n-Butylbenzene

Concen: 13.823 ug/l

RT: 14.035 min Scan# 2054

Delta R.T. 0.006 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :

VN1106MBS01





Tgt Ion: 91 Resp: 252713

Ion Ratio Lower Upper

91 100

92 52.2 25.9 77.8

134 23.5 11.7 35.0



#90

Hexachloroethane

Concen: 14.632 ug/l

RT: 14.312 min Scan# 2101

Delta R.T. -0.000 min

Lab File: VN088217.D




Acq: 06 Nov 2025 18:58


Tgt Ion: 117 Resp: 52298

Ion Ratio Lower Upper

117 100

201 60.6 29.1 87.4









#91
 1,2-Dichlorobenzene
 Concen: 17.597 ug/l
 RT: 14.082 min Scan# 2062
 Delta R.T. -0.000 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58




Instrument :
 MSVOA_N
 ClientSampleId :
 VN1106MBS01


Tgt Ion:146 Resp: 154129
 Ion Ratio Lower Upper
 146 100
 111 46.0 22.6 67.8
 148 62.0 31.9 95.9



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 18.406 ug/l
 RT: 14.700 min Scan# 2167
 Delta R.T. -0.000 min
 Lab File: VN088217.D
 Acq: 06 Nov 2025 18:58

Tgt Ion: 75 Resp: 28153
 Ion Ratio Lower Upper
 75 100
 155 65.6 36.6 109.9
 157 84.4 45.6 136.9









#93
1,2,4-Trichlorobenzene
Concen: 14.807 ug/l
RT: 15.364 min Scan# 21
Delta R.T. -0.000 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58




Instrument :
MSVOA_N
ClientSampleId :
VN1106MBSD01


Tgt Ion:180 Resp: 78547
Ion Ratio Lower Upper
180 100
182 94.7 47.6 142.9
145 32.5 16.5 49.5



#94
Hexachlorobutadiene
Concen: 12.297 ug/l
RT: 15.470 min Scan# 2298
Delta R.T. -0.006 min
Lab File: VN088217.D
Acq: 06 Nov 2025 18:58

Tgt Ion:225 Resp: 24292
Ion Ratio Lower Upper
225 100
223 63.2 32.0 96.2
227 61.3 32.1 96.5





#95

Naphthalene

Concen: 15.837 ug/l

RT: 15.617 min Scan# 21

Delta R.T. 0.006 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Instrument :

MSVOA_N

ClientSampleId :

VN1106MBS01





Tgt Ion:128 Resp: 287798

Ion Ratio Lower Upper

128 100

127 13.1 10.6 15.8

129 11.1 8.6 12.8



#96

1,2,3-Trichlorobenzene

Concen: 14.115 ug/l

RT: 15.811 min Scan# 2356

Delta R.T. -0.000 min

Lab File: VN088217.D

Acq: 06 Nov 2025 18:58

Tgt Ion:180 Resp: 71791

Ion Ratio Lower Upper

180 100

182 94.3 46.8 140.3

145 36.1 17.3 52.0

