

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110625\
 Data File : VN088205.D
 Acq On : 06 Nov 2025 14:49
 Operator : JC\MD
 Sample : Q3532-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Nov 07 02:19:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	226893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	513506	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	497289	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	239455	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	235809	60.102	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 120.200%		
35) Dibromofluoromethane	8.147	113	176907	51.281	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.560%		
50) Toluene-d8	10.547	98	668684	50.305	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.620%		
62) 4-Bromofluorobenzene	12.829	95	269048	57.312	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 114.620%		
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	39362	19.970	ug/l	98
20) Methylene Chloride	5.271	84	12749	3.853	ug/l #	89
25) 2-Butanone	7.477	43	45992	17.706	ug/l	94
43) Isopropyl Acetate	8.677	43	30566	3.060	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110625\
Data File : VN088205.D
Acq On : 06 Nov 2025 14:49
Operator : JC\MD
Sample : Q3532-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Nov 07 02:19:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

