

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110625\
 Data File : VN088206.D
 Acq On : 06 Nov 2025 15:10
 Operator : JC\MD
 Sample : Q3532-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Nov 07 02:20:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	224306	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	519367	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	506591	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	239334	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	237460	61.221	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 122.440%		
35) Dibromofluoromethane	8.153	113	176717	50.648	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.300%		
50) Toluene-d8	10.541	98	670928	49.905	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.800%		
62) 4-Bromofluorobenzene	12.823	95	266991	56.232	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 112.460%		
Target Compounds						
16) Acetone	4.430	43	41118	21.101	ug/l	98
20) Methylene Chloride	5.259	84	12158	3.717	ug/l #	96
25) 2-Butanone	7.471	43	46781	18.217	ug/l	100
43) Isopropyl Acetate	8.677	43	30171	2.987	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110625\
Data File : VN088206.D
Acq On : 06 Nov 2025 15:10
Operator : JC\MD
Sample : Q3532-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Nov 07 02:20:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

