

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110625\
 Data File : VN088208.D
 Acq On : 06 Nov 2025 15:52
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Nov 07 02:21:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	274999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	641595	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	626194	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	316853	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	308336	64.840	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	129.680%#
35) Dibromofluoromethane	8.147	113	225429	52.300	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.600%
50) Toluene-d8	10.547	98	823322	49.573	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.140%
62) 4-Bromofluorobenzene	12.829	95	335311	57.168	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	114.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110625\
Data File : VN088208.D
Acq On : 06 Nov 2025 15:52
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Nov 07 02:21:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

