

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110625\
 Data File : VN088214.D
 Acq On : 06 Nov 2025 17:56
 Operator : JC\MD
 Sample : Q3560-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SY-10I

Quant Time: Nov 07 02:23:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 03 01:16:06 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	219388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	502558	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	478183	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	233723	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	228398	60.205	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 120.400%	
35) Dibromofluoromethane	8.147	113	172195	51.002	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.000%	
50) Toluene-d8	10.541	98	647041	49.738	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.480%	
62) 4-Bromofluorobenzene	12.829	95	247299	53.827	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.660%	
Target Compounds						
						Qvalue
16) Acetone	4.424	43	5666	2.973	ug/l	94
30) Chloroform	7.947	83	75071	13.346	ug/l	99
64) Tetrachloroethene	11.088	164	1438	0.456	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110625\
Data File : VN088214.D
Acq On : 06 Nov 2025 17:56
Operator : JC\MD
Sample : Q3560-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SY-10I

Quant Time: Nov 07 02:23:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102325W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 03 01:16:06 2025
Response via : Initial Calibration

